

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2230 SAS No.: Q2230 SDG NO.: Q2230
Lab File ID: VX046635.D BFB Injection Date: 06/11/2025
Instrument ID: MSVOA_X BFB Injection Time: 14:01
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21
75	30.0 - 60.0% of mass 95	52.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	1.1 (1.6) 1
174	50.0 - 100.0% of mass 95	66.1
175	5.0 - 9.0% of mass 174	5 (7.5) 1
176	95.0 - 101.0% of mass 174	63.6 (96.3) 1
177	5.0 - 9.0% of mass 176	4.5 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046636.D	06/11/2025	14:46
VX0611WBL02	VX0611WBL02	VX046637.D	06/11/2025	15:14
VX0611WBS02	VX0611WBS02	VX046639.D	06/11/2025	15:58
GW-MW01-060425MS	Q2230-03MS	VX046641.D	06/11/2025	16:47
GW-MW01-060425MSD	Q2230-04MSD	VX046642.D	06/11/2025	17:09