

Preparation Blank Summary

Client: PARSONS Engineering of New York, Inc.

SDG No.: Q2231

Project: Con Ed Non MGP – Atlantic Ave 453957.600024.05

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136017BLW							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	06/05/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	06/05/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	06/05/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	06/05/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	06/05/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	06/05/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	06/05/2025
Sample ID: LB136041BL							
TDS	mg/L	< 5.0000	5.0000	U	1.0	10	06/06/2025