

Duplicate Sample Summary

Client: PARSONS Engineering of New York, Inc.	SDG No.: Q2231
Project: Con Ed Non MGP – Atlantic Ave 453957.600024.05	Sample ID: Q2231-04
Client ID: MW-16D-20250604MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Chloride	mg/L	+/-20	294	OR	295	OR	1	0		06/05/2025
Sulfate	mg/L	+/-20	68.9	OR	66.4	OR	1	4		06/05/2025
Nitrate	mg/L	+/-20	7.50	OR	7.00	OR	1	7		06/05/2025
Bromide	mg/L	+/-20	8.10		6.40		1	23	*	06/05/2025
Nitrite	mg/L	+/-20	2.30		1.80		1	24	*	06/05/2025
Orthophosphate as P	mg/L	+/-20	4.10		3.20		1	25	*	06/05/2025
Fluoride	mg/L	+/-20	1.70		1.30		1	27	*	06/05/2025

Duplicate Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2231
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Sample ID:	Q2231-05
Client ID:	MW-17-20250604DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	296		302		1	2.01		06/06/2025