

Laboratory Control Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2231
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Run No.:	LB136017

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136017BSW							
Bromide	mg/L	10	10.3		103	1	90-110	06/05/2025
Chloride	mg/L	3	3.10		103	1	90-110	06/05/2025
Fluoride	mg/L	2	2.00		100	1	90-110	06/05/2025
Nitrite	mg/L	3	3.10		103	1	90-110	06/05/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	06/05/2025
Sulfate	mg/L	15	15.2		101	1	90-110	06/05/2025
Orthophosphate as P	mg/L	5	5.00		100	1	90-110	06/05/2025

Laboratory Control Sample Summary

Client:	PARSONS Engineering of New York, Inc.	SDG No.:	Q2231
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Run No.:	LB136041

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136041BS							
TDS	mg/L	100	95.0		95	1	90-110	06/06/2025