

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2231 SAS No.: Q2231 SDG No.: Q2231

Instrument ID: BNA_M Calibration Date/Time: 06/10/2025 04:48

Lab File ID: BM050262.D Init. Calib. Date(s): 06/05/2025 06/05/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:20 13:56

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.199	1.232		2.8	
Benzaldehyde	0.995	0.967		-2.8	
Phenol-d6	1.578	1.600		1.4	
Phenol	1.670	1.704		2.0	20.0
bis(2-Chloroethyl)ether	1.343	1.373		2.2	
2-Chlorophenol	1.319	1.338		1.4	
2-Methylphenol	1.102	1.105		0.3	
2,2-oxybis(1-Chloropropane)	0.933	1.118		19.8	
Acetophenone	0.500	0.498		-0.4	
3+4-Methylphenols	1.474	1.472		-0.1	
n-Nitroso-di-n-propylamine	0.983	1.010	0.050	2.7	
Nitrobenzene-d5	0.385	0.405		5.2	
Hexachloroethane	0.555	0.579		4.3	
Nitrobenzene	0.350	0.366		4.6	
Isophorone	0.664	0.664		0.0	
2-Nitrophenol	0.151	0.156		3.3	20.0
2,4-Dimethylphenol	0.310	0.307		-1.0	
bis(2-Chloroethoxy)methane	0.434	0.449		3.5	
2,4-Dichlorophenol	0.287	0.284		-1.0	20.0
Naphthalene	1.034	1.006		-2.7	
4-Chloroaniline	0.441	0.433		-1.8	
Hexachlorobutadiene	0.190	0.181		-4.7	20.0
Caprolactam	0.091	0.088		-3.3	
4-Chloro-3-methylphenol	0.306	0.302		-1.3	20.0
2-Methylnaphthalene	0.624	0.605		-3.0	
Hexachlorocyclopentadiene	0.351	0.358	0.050	2.0	
2,4,6-Trichlorophenol	0.380	0.384		1.1	20.0
2-Fluorobiphenyl	1.477	1.442		-2.4	
2,4,5-Trichlorophenol	0.417	0.418		0.2	
1,1-Biphenyl	1.498	1.476		-1.5	
2-Chloronaphthalene	1.164	1.145		-1.6	
2-Nitroaniline	0.256	0.296		15.6	
Dimethylphthalate	1.353	1.338		-1.1	
Acenaphthylene	1.883	1.863		-1.1	
2,6-Dinitrotoluene	0.273	0.287		5.1	
3-Nitroaniline	0.304	0.322		5.9	
Acenaphthene	1.178	1.097		-6.9	20.0
2,4-Dinitrophenol	0.146	0.058	0.050	-60.3	
4-Nitrophenol	0.259	0.266	0.050	2.7	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2231 SAS No.: Q2231 SDG No.: Q2231

Instrument ID: BNA_M Calibration Date/Time: 06/10/2025 04:48

Lab File ID: BM050262.D Init. Calib. Date(s): 06/05/2025 06/05/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 09:20 13:56

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.723	1.686		-2.1	
2,4-Dinitrotoluene	0.360	0.377		4.7	
Diethylphthalate	1.342	1.344		0.1	
4-Chlorophenyl-phenylether	0.638	0.622		-2.5	
Fluorene	1.320	1.281		-3.0	
4-Nitroaniline	0.285	0.303		6.3	
4,6-Dinitro-2-methylphenol	0.109	0.068		-37.6	
n-Nitrosodiphenylamine	0.628	0.624		-0.6	20.0
2,4,6-Tribromophenol	0.227	0.228		0.4	
4-Bromophenyl-phenylether	0.213	0.212		-0.5	
Hexachlorobenzene	0.249	0.248		-0.4	
Atrazine	0.200	0.207		3.5	
Pentachlorophenol	0.162	0.224		38.3	20.0
Phenanthrene	1.143	1.117		-2.3	
Anthracene	1.143	1.125		-1.6	
Carbazole	1.041	1.038		-0.3	
Di-n-butylphthalate	1.133	1.226		8.2	
Fluoranthene	1.205	1.223		1.5	20.0
Pyrene	1.348	1.288		-4.5	
Terphenyl-d14	1.055	1.019		-3.4	
Butylbenzylphthalate	0.450	0.493		9.6	
3,3-Dichlorobenzidine	0.386	0.426		10.4	
Benzo (a) anthracene	1.266	1.246		-1.6	
Chrysene	1.204	1.189		-1.2	
Bis(2-ethylhexyl)phthalate	0.693	0.784		13.1	
Di-n-octyl phthalate	1.028	1.197		16.4	20.0
Benzo (b) fluoranthene	1.163	1.171		0.7	
Benzo (k) fluoranthene	1.187	1.165		-1.9	
Benzo (a) pyrene	1.093	1.115		2.0	20.0
Indeno (1,2,3-cd) pyrene	1.288	1.368		6.2	
Dibenzo (a,h) anthracene	1.045	1.119		7.1	
Benzo (g,h,i) perylene	1.046	1.121		7.2	
1,2,4,5-Tetrachlorobenzene	0.578	0.570		-1.4	
1,4-Dioxane	0.526	0.542		3.0	20.0
2,3,4,6-Tetrachlorophenol	0.361	0.354		-1.9	

All other compounds must meet a minimum RRF of 0.010.