

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2231 SAS No.: Q2231 SDG No.: Q2231

Instrument ID: BNA_P Calibration Date/Time: 06/09/2025 10:44

Lab File ID: BP024871.D Init. Calib. Date(s): 06/06/2025 06/06/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 15:18

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.198	1.175		-1.9	
Benzaldehyde	0.914	0.893		-2.3	
Phenol-d6	1.585	1.601		1.0	
Phenol	1.634	1.666		2.0	20.0
bis(2-Chloroethyl)ether	1.285	1.280		-0.4	
2-Chlorophenol	1.358	1.342		-1.2	
2-Methylphenol	1.141	1.152		1.0	
2,2-oxybis(1-Chloropropane)	1.682	1.631		-3.0	
Acetophenone	0.505	0.492		-2.6	
3+4-Methylphenols	1.557	1.575		1.2	
n-Nitroso-di-n-propylamine	1.078	1.088	0.050	0.9	
Nitrobenzene-d5	0.412	0.405		-1.7	
Hexachloroethane	0.576	0.542		-5.9	
Nitrobenzene	0.366	0.359		-1.9	
Isophorone	0.713	0.693		-2.8	
2-Nitrophenol	0.180	0.181		0.6	20.0
2,4-Dimethylphenol	0.309	0.303		-1.9	
bis(2-Chloroethoxy)methane	0.426	0.409		-4.0	
2,4-Dichlorophenol	0.296	0.296		0.0	20.0
Naphthalene	1.025	0.984		-4.0	
4-Chloroaniline	0.429	0.434		1.2	
Hexachlorobutadiene	0.201	0.187		-7.0	20.0
Caprolactam	0.109	0.112		2.8	
4-Chloro-3-methylphenol	0.342	0.347		1.5	20.0
2-Methylnaphthalene	0.650	0.642		-1.2	
Hexachlorocyclopentadiene	0.346	0.332	0.050	-4.0	
2,4,6-Trichlorophenol	0.381	0.373		-2.1	20.0
2-Fluorobiphenyl	1.485	1.394		-6.1	
2,4,5-Trichlorophenol	0.408	0.416		2.0	
1,1-Biphenyl	1.453	1.381		-5.0	
2-Chloronaphthalene	1.117	1.069		-4.3	
2-Nitroaniline	0.343	0.351		2.3	
Dimethylphthalate	1.475	1.400		-5.1	
Acenaphthylene	1.863	1.770		-5.0	
2,6-Dinitrotoluene	0.318	0.308		-3.1	
3-Nitroaniline	0.330	0.343		3.9	
Acenaphthene	1.067	1.029		-3.6	20.0
2,4-Dinitrophenol	0.178	0.182	0.050	2.2	
4-Nitrophenol	0.239	0.252	0.050	5.4	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2231 SAS No.: Q2231 SDG No.: Q2231

Instrument ID: BNA_P Calibration Date/Time: 06/09/2025 10:44

Lab File ID: BP024871.D Init. Calib. Date(s): 06/06/2025 06/06/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:30 15:18

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.713	1.630		-4.8	
2,4-Dinitrotoluene	0.445	0.443		-0.4	
Diethylphthalate	1.470	1.395		-5.1	
4-Chlorophenyl-phenylether	0.677	0.627		-7.4	
Fluorene	1.384	1.326		-4.2	
4-Nitroaniline	0.299	0.327		9.4	
4,6-Dinitro-2-methylphenol	0.131	0.131		0.0	
n-Nitrosodiphenylamine	0.620	0.596		-3.9	20.0
2,4,6-Tribromophenol	0.277	0.273		-1.4	
4-Bromophenyl-phenylether	0.226	0.216		-4.4	
Hexachlorobenzene	0.274	0.261		-4.7	
Atrazine	0.225	0.214		-4.9	
Pentachlorophenol	0.142	0.150		5.6	20.0
Phenanthrene	1.105	1.056		-4.4	
Anthracene	1.119	1.067		-4.6	
Carbazole	1.038	1.005		-3.2	
Di-n-butylphthalate	1.285	1.195		-7.0	
Fluoranthene	1.281	1.200		-6.3	20.0
Pyrene	1.249	1.182		-5.4	
Terphenyl-d14	1.116	1.046		-6.3	
Butylbenzylphthalate	0.572	0.544		-4.9	
3,3-Dichlorobenzidine	0.508	0.491		-3.3	
Benzo (a) anthracene	1.279	1.229		-3.9	
Chrysene	1.212	1.154		-4.8	
Bis(2-ethylhexyl)phthalate	0.820	0.772		-5.9	
Di-n-octyl phthalate	1.445	1.351		-6.5	20.0
Benzo (b) fluoranthene	1.145	1.096		-4.3	
Benzo (k) fluoranthene	1.165	1.086		-6.8	
Benzo (a) pyrene	1.118	1.057		-5.5	20.0
Indeno (1,2,3-cd) pyrene	1.459	1.436		-1.6	
Dibenzo (a,h) anthracene	1.188	1.170		-1.5	
Benzo (g,h,i) perylene	1.179	1.158		-1.8	
1,2,4,5-Tetrachlorobenzene	0.568	0.541		-4.8	
1,4-Dioxane	0.527	0.520		-1.3	20.0
2,3,4,6-Tetrachlorophenol	0.365	0.364		-0.3	

All other compounds must meet a minimum RRF of 0.010.