

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2231 SAS No.: Q2231 SDG NO.: Q2231

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/10/2025

Lab File ID: BM050262.D Time Analyzed: 04:48

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	460715	7.763	1877130	10.55	1070950	14.40
UPPER LIMIT	921430	8.263	3754260	11.051	2141900	14.898
LOWER LIMIT	230358	7.263	938565	10.051	535475	13.898
EPA SAMPLE NO.						
01 MW-14-20250604DL	306070	7.76	1203870	10.55	655906	14.39
02 MW-14-20250604DL2	323093	7.76	1225980	10.55	643043	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.