

# DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Laboratory Name : Alliance Technical Group LLC

Client :

JACOBS Engineering Group, Inc.

Project Location : Princeton Junction

Project Number : D3868221

Laboratory Sample ID(s) : Q2233

Sampling Date(s) : 06/04/2025

List DKQP Methods Used (e.g., 8260,8270, et Cetra)

**8260D**

|    |                                                                                                                                                                                                                                                                                                     |                                                                                                                                                                             |
|----|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1  | For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards? | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No                                                                                                         |
| 1A | Were the method specified handling, preservation, and holding time requirements met?                                                                                                                                                                                                                | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No                                                                                                         |
| 1B | EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)                                                                                                                                                                             | <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A                                                                            |
| 2  | Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?                                                                                                                                                           | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No                                                                                                         |
| 3  | Were samples received at an appropriate temperature (4±2° C)?                                                                                                                                                                                                                                       | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A                                                                            |
| 4  | Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?                                                                                                                                                                                                                 | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No                                                                                                         |
| 5  | a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt?<br><br>b)Were these reporting limits met?                                                                                                                        | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No<br><br><input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A |
| 6  | For each analytical method referenced in this laboratory report package, were results reported for all constituents identified in the method-specific analyte lists presented in the DKQP documents and/or site-specific QAPP?                                                                      | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No                                                                                                         |
| 7  | Are project-specific matrix spikes and/or laboratory duplicates included in this data set?                                                                                                                                                                                                          | <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No                                                                                                         |

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."

## Cover Page

**Order ID :** Q2233

**Project ID :** Former Schlumberger STC PTC Site D3868221

**Client :** JACOBS Engineering Group, Inc.

**Lab Sample Number**

Q2233-01  
Q2233-03  
Q2233-04

**Client Sample Number**

MW-18B-56-060425  
MW-18B-56-060425-FD  
MW-19B-72-060425

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : \_\_\_\_\_

Date: 6/17/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**JACOBS Engineering Group, Inc.**

**Project Name: Former Schlumberger STC PTC Site D3868221**

**Project # N/A**

**Order ID # Q2233**

**Test Name: VOCMS Group3**

### **A. Number of Samples and Date of Receipt:**

5 Water samples were received on 06/05/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: VOC-TRACE-SFAM and VOCMS Group3. This data package contains results for VOCMS Group3.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOCMS Group3 was based on method 8260D.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Samples MW-19B-72-060425 was diluted at straight dilution after checking past history of this sample.

Samples MW-18B-56-060425, MW-18B-56-060425-FD and MW-19B-72-060425 were diluted due to high concentrations.

### **E. Additional Comments:**

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

The SIM analysis is not required for the sample MW-18B-56-060425-SIM and MW-19B-72-060425-SIM as all the SIM target analytes are detected at or above the sample adjusted CRQLs in the full scan analysis, a SIM analysis is not to be performed for that sample."

Trip Blank was not provided with this set of samples.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_

## DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

|       |                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                                           |
|-------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Value | If the result is a value greater than or equal to the detection limit, report the value                                                                                                                                                                                                                                                                                                                                                                                                                                                   |
| U     | Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.                                                                                                                                                                                                                                 |
| ND    | Indicates the analyte was analyzed for, but not detected                                                                                                                                                                                                                                                                                                                                                                                                                                                                                  |
| J     | Indicates an estimated value. This flag is used:<br>(1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.)<br>(2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others. |
| B     | Indicates the analyte was found in the blank as well as the sample report as "12 B".                                                                                                                                                                                                                                                                                                                                                                                                                                                      |
| E     | Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.                                                                                                                                                                                                                                                                                                                                                                                                                         |
| D     | This flag identifies all compounds identified in an analysis at a secondary dilution factor.                                                                                                                                                                                                                                                                                                                                                                                                                                              |
| P     | This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".                                                                                                                                                                                                                                                                                                                        |
| N     | This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.                                                                                                                                                                                                                  |
| A     | This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.                                                                                                                                                                                                                                                                                                                                                                                                                                     |
| Q     | Indicates the LCS did not meet the control limits requirements                                                                                                                                                                                                                                                                                                                                                                                                                                                                            |

**APPENDIX A**

**QA REVIEW GENERAL DOCUMENTATION**

Project #: Q2233

Completed

For thorough review, the report must have the following:

**GENERAL:**

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

**COVER PAGE:**

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

**CHAIN OF CUSTODY:**

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

**ANALYTICAL:**

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 06/17/2025