

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2233

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Datafile : VX046590.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0610WBSD01	Vinyl chloride	20	19.1	ug/L	96	2		70 (65)	130 (117)	20 (19)
	1,1-Dichloroethene	20	19.8	ug/L	99	3		70 (74)	130 (110)	20 (20)
	1,1-Dichloroethane	20	20.7	ug/L	104	2		70 (78)	130 (112)	20 (20)
	cis-1,2-Dichloroethene	20	20.5	ug/L	103	0		70 (77)	130 (110)	20 (20)
	1,1,1-Trichloroethane	20	20.9	ug/L	104	1		70 (80)	130 (108)	20 (20)
	Benzene	20	20.6	ug/L	103	2		70 (82)	130 (109)	20 (15)
	1,2-Dichloroethane	20	20.6	ug/L	103	1		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.8	ug/L	99	2		70 (77)	130 (113)	20 (15)
	1,1,2-Trichloroethane	20	22.3	ug/L	112	4		70 (83)	130 (112)	20 (20)
	Tetrachloroethene	20	19.8	ug/L	99	3		70 (67)	130 (123)	20 (15)