

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05  
 Lab Code: CHEM Case No.: Q2233 SAS No.: Q2233 SDG NO.: Q2233  
 Lab File ID: VX046586.D Date Analyzed: 06/10/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:07  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                       | IS4<br>AREA # | RT #   |  |  |  |  |
|-----------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD           | 66941         | 12.018 |  |  |  |  |
| UPPER LIMIT           | 133882        | 12.518 |  |  |  |  |
| LOWER LIMIT           | 33470.5       | 11.518 |  |  |  |  |
| EPA SAMPLE NO.        |               |        |  |  |  |  |
| MW-18B-56-060425      | 116937        | 12.02  |  |  |  |  |
| MW-18B-56-060425DL    | 112325        | 12.02  |  |  |  |  |
| MW-18B-56-060425-FD   | 110562        | 12.02  |  |  |  |  |
| MW-18B-56-060425-FDDL | 113693        | 12.02  |  |  |  |  |
| MW-19B-72-060425      | 113207        | 12.02  |  |  |  |  |
| MW-19B-72-060425DL    | 116361        | 12.02  |  |  |  |  |
| VX0610WBL01           | 104374        | 12.02  |  |  |  |  |
| VX0610WBS01           | 68222         | 12.02  |  |  |  |  |
| VX0610WBSD01          | 65284         | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.