

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q2234
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q2234 **SAS No.:** Q2234
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	2000	2000	100	90 - 110	P	06/11/2025	18:48	LB136121
	Silver	46.7	50.0	93	90 - 110	P	06/11/2025	18:48	LB136121

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q2234

Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q2234 **SAS No.:** Q2234

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	52.6	50.0	105	80 - 120	P	06/11/2025	19:00	LB136121
	Silver	1.07	1.0	107	80 - 120	P	06/11/2025	19:00	LB136121

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. SDG No.: Q2234

Contract: JACO05 Lab Code: CHEM Case No.: Q2234 SAS No.: Q2234

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	127000	125000	102	90 - 110	P	06/11/2025	19:29	LB136121
	Silver	536	500	107	90 - 110	P	06/11/2025	19:29	LB136121
CCV02	Iron	130000	125000	104	90 - 110	P	06/11/2025	19:52	LB136121
	Silver	515	500	103	90 - 110	P	06/11/2025	19:52	LB136121
CCV03	Iron	129000	125000	103	90 - 110	P	06/11/2025	20:35	LB136121
	Silver	505	500	101	90 - 110	P	06/11/2025	20:35	LB136121
CCV04	Iron	128000	125000	103	90 - 110	P	06/11/2025	21:16	LB136121
	Silver	509	500	102	90 - 110	P	06/11/2025	21:16	LB136121
CCV05	Iron	128000	125000	103	90 - 110	P	06/11/2025	21:34	LB136121
	Silver	525	500	105	90 - 110	P	06/11/2025	21:34	LB136121