

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q2234
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q2234 **SAS No.:** Q2234
ICS Source: EPA **Instrument ID:** P7

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	103000	100000	103	0	0	06/11/2025	19:10	LB136121
	Silver	0.050			-2	2	06/11/2025	19:10	LB136121
ICSAB01	Iron	102000	100000	102	0	0	06/11/2025	19:13	LB136121
	Silver	20.3	18.0	113	15.3	20.7	06/11/2025	19:13	LB136121