

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2234 SAS No.: Q2234 SDG NO.: Q2234

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/09/2025

Lab File ID: BN037204.D Time Analyzed: 23:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1688	7.589	4447	10.37	2457	14.23
UPPER LIMIT	3376	8.089	8894	10.872	4914	14.734
LOWER LIMIT	844	7.089	2223.5	9.872	1228.5	13.734
EPA SAMPLE NO.						
01 MW-18B-56-060425	1612	7.59	4076	10.36	2065	14.24
02 MW-18B-56-060425-FD	1777	7.59	4601	10.36	2273	14.23
03 MW-19B-72-060425	2075	7.59	5375	10.36	3061	14.23
04 MW-19B-72-060425DL	1883	7.58	4807	10.36	2597	14.24
05 MW-17B-55-060425	1414	7.59	3747	10.37	2035	14.24

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.