

Instrument ID: MSVOA\_X

**Daily Analysis Runlog For Sequence/QC Batch ID # VX060625**

|              |                     |                   |                                   |
|--------------|---------------------|-------------------|-----------------------------------|
| Review By    | Mahesh Dadoda       | Review On         | 6/9/2025 1:15:21 PM               |
| Supervise By | Semsettin Yesilyurt | Supervise On      | 6/9/2025 1:19:58 PM               |
| SubDirectory | VX060625            | HP Acquire Method | HP Processing Method 82X060625W.M |

| STD. NAME                                                          | STD REF.#                                                         |
|--------------------------------------------------------------------|-------------------------------------------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                            | VP134153<br>VP134235,VP134236,VP134237,VP134238,VP134239,VP134240 |
| CCC<br>Internal Standard/PEM                                       | VP134154                                                          |
| ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134241                                                          |

| Sr# | SampleID     | ClientID     | Data File Name | Date-Time         | Comment       | Operator | Status |
|-----|--------------|--------------|----------------|-------------------|---------------|----------|--------|
| 1   | BFB          | BFB          | VX046516.D     | 06 Jun 2025 08:47 |               | JC/MD    | Ok     |
| 2   | VSTDICC001   | VSTDICC001   | VX046517.D     | 06 Jun 2025 09:13 |               | JC/MD    | Not Ok |
| 3   | VSTDICC005   | VSTDICC005   | VX046518.D     | 06 Jun 2025 09:42 | for TCLP      | JC/MD    | Ok,M   |
| 4   | VSTDICC020   | VSTDICC020   | VX046519.D     | 06 Jun 2025 10:18 | Comp #05 fail | JC/MD    | Ok,M   |
| 5   | VSTDICCC050  | VSTDICCC050  | VX046520.D     | 06 Jun 2025 10:40 | LR-13,16,17   | JC/MD    | Ok,M   |
| 6   | VSTDICC100   | VSTDICC100   | VX046521.D     | 06 Jun 2025 11:02 |               | JC/MD    | Ok,M   |
| 7   | VSTDICC150   | VSTDICC150   | VX046522.D     | 06 Jun 2025 11:25 |               | JC/MD    | Ok,M   |
| 8   | IBLK         | IBLK         | VX046523.D     | 06 Jun 2025 11:47 |               | JC/MD    | Ok     |
| 9   | VSTDICC001   | VSTDICC001   | VX046524.D     | 06 Jun 2025 12:57 |               | JC/MD    | Ok,M   |
| 10  | VSTDICV050   | ICV VX060625 | VX046525.D     | 06 Jun 2025 14:11 |               | JC/MD    | Ok,M   |
| 11  | VX0606MBL01  | VX0606MBL01  | VX046526.D     | 06 Jun 2025 14:39 |               | JC/MD    | Ok     |
| 12  | VX0606WBL01  | VX0606WBL01  | VX046527.D     | 06 Jun 2025 15:02 |               | JC/MD    | Ok     |
| 13  | VX0606WBS01  | VX0606WBS01  | VX046528.D     | 06 Jun 2025 15:25 |               | JC/MD    | Ok,M   |
| 14  | VX0606MBS01  | VX0606MBS01  | VX046529.D     | 06 Jun 2025 15:51 |               | JC/MD    | Ok,M   |
| 15  | Q2168-11MEDL | C2MEDL       | VX046530.D     | 06 Jun 2025 16:13 |               | JC/MD    | Ok     |
| 16  | VX0606WBSD01 | VX0606WBSD01 | VX046531.D     | 06 Jun 2025 16:36 |               | JC/MD    | Ok,M   |
| 17  | Q2194-02     | COMP-12      | VX046532.D     | 06 Jun 2025 16:58 | vial A pH#5.0 | JC/MD    | Ok     |
| 18  | Q2194-04     | COMP-13      | VX046533.D     | 06 Jun 2025 17:21 | vial A pH#5.0 | JC/MD    | Ok     |

Instrument ID: MSVOA\_X

**Daily Analysis Runlog For Sequence/QC Batch ID # VX060625**

| Review By                                                          | Maresh Dadoda                                                     | Review On         | 6/9/2025 1:15:21 PM               |
|--------------------------------------------------------------------|-------------------------------------------------------------------|-------------------|-----------------------------------|
| Supervise By                                                       | Semsettin Yesilyurt                                               | Supervise On      | 6/9/2025 1:19:58 PM               |
| SubDirectory                                                       | VX060625                                                          | HP Acquire Method | HP Processing Method 82X060625W.M |
| STD. NAME                                                          | STD REF.#                                                         |                   |                                   |
| Tune/Reschk<br>Initial Calibration Stds                            | VP134153<br>VP134235,VP134236,VP134237,VP134238,VP134239,VP134240 |                   |                                   |
| CCC<br>Internal Standard/PEM                                       | VP134154                                                          |                   |                                   |
| ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134241                                                          |                   |                                   |

|    |              |                |            |                   |                           |       |        |
|----|--------------|----------------|------------|-------------------|---------------------------|-------|--------|
| 19 | Q2207-09     | BU-703-COMP-01 | VX046534.D | 06 Jun 2025 17:43 | vial A pH#5.0             | JC/MD | Ok,M   |
| 20 | Q2207-18     | BU-703-COMP-02 | VX046535.D | 06 Jun 2025 18:06 | vial A pH#5.0             | JC/MD | Ok,M   |
| 21 | Q2207-27     | BU-703-COMP-03 | VX046536.D | 06 Jun 2025 18:28 | vial A pH#5.0             | JC/MD | Ok,M   |
| 22 | Q2207-36     | BU-703-COMP-04 | VX046537.D | 06 Jun 2025 18:51 | vial A pH#5.0             | JC/MD | Ok,M   |
| 23 | Q2207-45     | BU-703-COMP-05 | VX046538.D | 06 Jun 2025 19:13 | vial A pH#5.0             | JC/MD | Ok,M   |
| 24 | Q2208-09     | BU-703-COMP-06 | VX046539.D | 06 Jun 2025 19:35 | vial A pH#5.0             | JC/MD | Ok,M   |
| 25 | Q2208-18     | BU-703-COMP-07 | VX046540.D | 06 Jun 2025 19:58 | vial A pH#5.0             | JC/MD | Ok,M   |
| 26 | Q2208-27     | BU-703-COMP-08 | VX046541.D | 06 Jun 2025 20:20 | vial A pH#5.0             | JC/MD | Ok,M   |
| 27 | Q2208-36     | BU-703-COMP-09 | VX046542.D | 06 Jun 2025 20:42 | vial A pH#5.0             | JC/MD | Ok,M   |
| 28 | Q2236-01     | WC-A4-05A-G    | VX046543.D | 06 Jun 2025 21:04 | vial A pH#5.0 Out of tune | JC/MD | Not Ok |
| 29 | VSTDCCC050   | VSTDCCC050EC   | VX046544.D | 06 Jun 2025 21:26 | Out of tune               | JC/MD | Not Ok |
| 30 | BFB          | BFB            | VX046545.D | 06 Jun 2025 23:59 |                           | JC/MD | Ok     |
| 31 | VSTDCCC050   | VSTDCCC050     | VX046546.D | 07 Jun 2025 00:34 |                           | JC/MD | Ok,M   |
| 32 | VX0606WBL02  | VX0606WBL02    | VX046547.D | 07 Jun 2025 01:17 |                           | JC/MD | Ok     |
| 33 | VX0606WBS02  | VX0606WBS02    | VX046548.D | 07 Jun 2025 02:00 |                           | JC/MD | Ok,M   |
| 34 | VX0606WBSD02 | VX0606WBSD02   | VX046549.D | 07 Jun 2025 02:22 |                           | JC/MD | Ok,M   |
| 35 | PB168312TB   | PB168312TB     | VX046550.D | 07 Jun 2025 02:43 |                           | JC/MD | Ok,M   |
| 36 | PB168272TB   | PB168272TB     | VX046551.D | 07 Jun 2025 03:05 |                           | JC/MD | Ok,M   |
| 37 | Q2236-05     | WC-A2-04-G     | VX046552.D | 07 Jun 2025 03:26 | Surrogate Fail            | JC/MD | ReRun  |

Instrument ID: MSVOA\_X

**Daily Analysis Runlog For Sequence/QC Batch ID # VX060625**

|                                                                    |                                                                   |                   |                                   |
|--------------------------------------------------------------------|-------------------------------------------------------------------|-------------------|-----------------------------------|
| Review By                                                          | Maresh Dadoda                                                     | Review On         | 6/9/2025 1:15:21 PM               |
| Supervise By                                                       | Semsettin Yesilyurt                                               | Supervise On      | 6/9/2025 1:19:58 PM               |
| SubDirectory                                                       | VX060625                                                          | HP Acquire Method | HP Processing Method 82X060625W.M |
| <b>STD. NAME</b>                                                   | <b>STD REF.#</b>                                                  |                   |                                   |
| Tune/Reschk<br>Initial Calibration Stds                            | VP134153<br>VP134235,VP134236,VP134237,VP134238,VP134239,VP134240 |                   |                                   |
| CCC<br>Internal Standard/PEM                                       | VP134154                                                          |                   |                                   |
| ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134241                                                          |                   |                                   |

|    |            |              |            |                   |                                        |       |       |
|----|------------|--------------|------------|-------------------|----------------------------------------|-------|-------|
| 38 | Q2236-09   | WC-A2-05-G   | VX046553.D | 07 Jun 2025 03:48 | Internal Standard Fail                 | JC/MD | ReRun |
| 39 | Q2236-13   | WC-A2-06-G   | VX046554.D | 07 Jun 2025 04:09 | vial A pH#5.0                          | JC/MD | Ok    |
| 40 | Q2236-17   | WC-A2-07-G   | VX046555.D | 07 Jun 2025 04:31 | Internal Standard Fail; Surrogate fail | JC/MD | ReRun |
| 41 | Q2227-04   | TP07-MHH-WC  | VX046556.D | 07 Jun 2025 04:52 |                                        | JC/MD | Ok,M  |
| 42 | Q2228-04   | TP08-MHI-WC  | VX046557.D | 07 Jun 2025 05:14 | Internal Standard Fail                 | JC/MD | ReRun |
| 43 | Q2235-01   | WC-A2-08-G   | VX046558.D | 07 Jun 2025 05:36 | Internal Standard Fail                 | JC/MD | ReRun |
| 44 | Q2240-04   | TP-3         | VX046559.D | 07 Jun 2025 05:57 | Internal Standard Fail                 | JC/MD | ReRun |
| 45 | Q2240-08   | TP-2         | VX046560.D | 07 Jun 2025 06:18 | Internal Standard Fail                 | JC/MD | ReRun |
| 46 | Q2240-12   | TP-1         | VX046561.D | 07 Jun 2025 06:40 | Internal Standard Fail                 | JC/MD | ReRun |
| 47 | Q2241-04   | TP-N         | VX046562.D | 07 Jun 2025 07:01 | Internal Standard Fail                 | JC/MD | ReRun |
| 48 | Q2241-08   | TP-S         | VX046563.D | 07 Jun 2025 07:23 | Internal Standard Fail                 | JC/MD | ReRun |
| 49 | Q2226-04   | TP06-MHI-WC  | VX046564.D | 07 Jun 2025 07:44 | Internal Standard Fail                 | JC/MD | ReRun |
| 50 | VSTDCCC050 | VSTDCCC050EC | VX046565.D | 07 Jun 2025 08:06 |                                        | JC/MD | Ok,M  |

M : Manual Integration

Instrument ID: MSVOA\_X

**Daily Analysis Runlog For Sequence/QC Batch ID # VX061025**

|              |               |                   |                                   |
|--------------|---------------|-------------------|-----------------------------------|
| Review By    | John Carlone  | Review On         | 6/11/2025 10:17:47 AM             |
| Supervise By | Maresh Dadoda | Supervise On      | 6/11/2025 11:14:33 AM             |
| SubDirectory | VX061025      | HP Acquire Method | HP Processing Method 82X060625W.M |

| STD. NAME                                                                                          | STD REF.#         |
|----------------------------------------------------------------------------------------------------|-------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134200          |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134201,VP134202 |

| Sr# | SampleID     | ClientID            | Data File Name | Date-Time         | Comment        | Operator | Status |
|-----|--------------|---------------------|----------------|-------------------|----------------|----------|--------|
| 1   | BFB          | BFB                 | VX046585.D     | 10 Jun 2025 08:36 |                | JC/MD    | Ok     |
| 2   | VSTDCCC050   | VSTDCCC050          | VX046586.D     | 10 Jun 2025 09:07 | pH#Lot#V12668  | JC/MD    | Ok,M   |
| 3   | VX0610MBL01  | VX0610MBL01         | VX046587.D     | 10 Jun 2025 09:36 |                | JC/MD    | Ok     |
| 4   | VX0610WBL01  | VX0610WBL01         | VX046588.D     | 10 Jun 2025 09:57 |                | JC/MD    | Ok     |
| 5   | VX0610WBS01  | VX0610WBS01         | VX046589.D     | 10 Jun 2025 10:18 |                | JC/MD    | Ok,M   |
| 6   | VX0610WBSD01 | VX0610WBSD01        | VX046590.D     | 10 Jun 2025 10:44 |                | JC/MD    | Ok,M   |
| 7   | Q2262-02     | ARS20-0032          | VX046591.D     | 10 Jun 2025 11:05 | vial B pH#5.0  | JC/MD    | Ok     |
| 8   | Q2262-04     | ARS20-0001          | VX046592.D     | 10 Jun 2025 11:27 | vial B pH#5.0  | JC/MD    | Ok     |
| 9   | IBLK         | IBLK                | VX046593.D     | 10 Jun 2025 11:48 |                | JC/MD    | Ok     |
| 10  | Q2230-01     | FB-060425           | VX046594.D     | 10 Jun 2025 12:09 | vial A pH<2 FB | JC/MD    | Ok     |
| 11  | Q2230-06     | TB060425            | VX046595.D     | 10 Jun 2025 12:30 | vial A pH<2 TB | JC/MD    | Ok     |
| 12  | Q2233-01DL   | MW-18B-56-060425DL  | VX046596.D     | 10 Jun 2025 12:51 | vial B pH<2    | JC/MD    | Ok     |
| 13  | Q2233-03DL   | MW-18B-56-060425-FD | VX046597.D     | 10 Jun 2025 13:13 | vial B pH<2    | JC/MD    | Ok     |
| 14  | Q2233-04DL   | MW-19B-72-060425DL  | VX046598.D     | 10 Jun 2025 13:34 | vial B pH<2    | JC/MD    | Ok     |
| 15  | Q2234-01DL   | MW-17B-55-060425DL  | VX046599.D     | 10 Jun 2025 13:55 | vial B pH<2    | JC/MD    | Ok     |
| 16  | IBLK         | IBLK                | VX046600.D     | 10 Jun 2025 14:16 |                | JC/MD    | Ok     |
| 17  | Q2230-05     | GW-MW901-060425     | VX046601.D     | 10 Jun 2025 14:38 | vial A pH<2    | JC/MD    | Ok,M   |
| 18  | Q2230-02     | GW-MW01-060425      | VX046602.D     | 10 Jun 2025 14:59 | vial A pH<2    | JC/MD    | Ok,M   |

Instrument ID: MSVOA\_X

**Daily Analysis Runlog For Sequence/QC Batch ID # VX061025**

|                                                                                                    |                   |                   |                                   |
|----------------------------------------------------------------------------------------------------|-------------------|-------------------|-----------------------------------|
| Review By                                                                                          | John Carlone      | Review On         | 6/11/2025 10:17:47 AM             |
| Supervise By                                                                                       | Maresh Dadoda     | Supervise On      | 6/11/2025 11:14:33 AM             |
| SubDirectory                                                                                       | VX061025          | HP Acquire Method | HP Processing Method 82X060625W.M |
| <b>STD. NAME</b>                                                                                   | <b>STD REF.#</b>  |                   |                                   |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134200          |                   |                                   |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134201,VP134202 |                   |                                   |

|    |             |                     |            |                   |                                                         |       |          |
|----|-------------|---------------------|------------|-------------------|---------------------------------------------------------|-------|----------|
| 19 | Q2230-03MS  | GW-MW01-060425MS    | VX046603.D | 10 Jun 2025 15:20 | Internal Standard Fail; Surrogate Fail; Spike not added | JC/MD | Not Ok   |
| 20 | Q2230-04MSD | GW-MW01-060425MS    | VX046604.D | 10 Jun 2025 15:42 | Internal Standard Fail; Surrogate Fail; Spike not added | JC/MD | Not Ok   |
| 21 | IBLK        | IBLK                | VX046605.D | 10 Jun 2025 16:03 |                                                         | JC/MD | Ok       |
| 22 | Q2249-01    | MW-06-6.5-060525    | VX046606.D | 10 Jun 2025 16:24 | Run @ lower dilution                                    | JC/MD | Not Ok   |
| 23 | IBLK        | IBLK                | VX046607.D | 10 Jun 2025 16:46 |                                                         | JC/MD | Ok       |
| 24 | Q2233-01    | MW-18B-56-060425    | VX046608.D | 10 Jun 2025 17:07 | vial A pH<2 Need 10X                                    | JC/MD | Dilution |
| 25 | Q2233-03    | MW-18B-56-060425-FD | VX046609.D | 10 Jun 2025 17:28 | vial A pH<2 Need 10X                                    | JC/MD | Dilution |
| 26 | Q2233-04    | MW-19B-72-060425    | VX046610.D | 10 Jun 2025 17:50 | vial A pH<2 Need 100X                                   | JC/MD | Dilution |
| 27 | Q2234-01    | MW-17B-55-060425    | VX046611.D | 10 Jun 2025 18:11 | vial A pH<2 Need 100X                                   | JC/MD | Dilution |
| 28 | VSTDCCC050  | VSTDCCC050EC        | VX046612.D | 10 Jun 2025 18:53 |                                                         | JC/MD | Ok,M     |
| 29 | BFB         | BFB                 | VX046613.D | 10 Jun 2025 19:57 |                                                         | JC/MD | Ok       |
| 30 | VSTDICC005  | VSTDICC005          | VX046614.D | 10 Jun 2025 20:19 |                                                         | JC/MD | Ok,M     |
| 31 | VSTDICCC020 | VSTDICCC020         | VX046615.D | 10 Jun 2025 20:40 |                                                         | JC/MD | Ok,M     |
| 32 | VSTDICC050  | VSTDICC050          | VX046616.D | 10 Jun 2025 21:01 |                                                         | JC/MD | Ok,M     |
| 33 | VSTDICC100  | VSTDICC100          | VX046617.D | 10 Jun 2025 21:22 |                                                         | JC/MD | Ok,M     |
| 34 | VSTDICC150  | VSTDICC150          | VX046618.D | 10 Jun 2025 21:44 |                                                         | JC/MD | Ok,M     |
| 35 | IBLK        | IBLK                | VX046619.D | 10 Jun 2025 22:05 |                                                         | JC/MD | Ok       |
| 36 | VSTDICV020  | ICV VX061025        | VX046620.D | 10 Jun 2025 22:26 |                                                         | JC/MD | Ok,M     |
| 37 | VX0610WBS02 | VX0610WBS02         | VX046621.D | 10 Jun 2025 23:09 |                                                         | JC/MD | Ok,M     |

Instrument ID: MSVOA\_X

**Daily Analysis Runlog For Sequence/QC Batch ID # VX061025**

|                                                                                                    |                   |                   |                                   |
|----------------------------------------------------------------------------------------------------|-------------------|-------------------|-----------------------------------|
| Review By                                                                                          | John Carlone      | Review On         | 6/11/2025 10:17:47 AM             |
| Supervise By                                                                                       | Maresh Dadoda     | Supervise On      | 6/11/2025 11:14:33 AM             |
| SubDirectory                                                                                       | VX061025          | HP Acquire Method | HP Processing Method 82X060625W.M |
| <b>STD. NAME</b>                                                                                   | <b>STD REF.#</b>  |                   |                                   |
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134200          |                   |                                   |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134201,VP134202 |                   |                                   |

|    |              |                     |            |                   |                     |       |          |
|----|--------------|---------------------|------------|-------------------|---------------------|-------|----------|
| 38 | VX0610WBSD02 | VX0610WBSD02        | VX046622.D | 10 Jun 2025 23:30 |                     | JC/MD | Ok,M     |
| 39 | VX0610WBL02  | VX0610WBL02         | VX046623.D | 11 Jun 2025 00:12 |                     | JC/MD | Ok       |
| 40 | Q2203-01     | 001-WILLETS-PT-BLVD | VX046624.D | 11 Jun 2025 00:33 | vial A pH<2 Need 5X | JC/MD | Dilution |
| 41 | Q2203-04     | 002-35TH-AVE(JUNE)  | VX046625.D | 11 Jun 2025 00:54 | vial A pH<2 Need 5X | JC/MD | Dilution |
| 42 | VSTDCCC020   | VSTDCCC020EC        | VX046626.D | 11 Jun 2025 01:16 |                     | JC/MD | Ok,M     |

M : Manual Integration