

Surrogate Summary

SDG No.: Q2234

Client: JACOBS Engineering Group, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2234-01	MW-17B-55-060425	1,2-Dichloroethane-d4	50	50.7	101	70 (74)	130 (125)
		Dibromofluoromethane	50	50.0	100	70 (75)	130 (124)
		Toluene-d8	50	53.7	107	70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.8	112	70 (77)	130 (121)
Q2234-01DL	MW-17B-55-060425DL	1,2-Dichloroethane-d4	50	50.4	101	70 (74)	130 (125)
		Dibromofluoromethane	50	50.2	100	70 (75)	130 (124)
		Toluene-d8	50	54.4	109	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.7	109	70 (77)	130 (121)
VX0610WBL01	VX0610WBL01	1,2-Dichloroethane-d4	50	50.1	100	70 (74)	130 (125)
		Dibromofluoromethane	50	50.1	100	70 (75)	130 (124)
		Toluene-d8	50	54.3	109	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.8	110	70 (77)	130 (121)
VX0610WBS01	VX0610WBS01	1,2-Dichloroethane-d4	50	47.4	95	70 (74)	130 (125)
		Dibromofluoromethane	50	53.0	106	70 (75)	130 (124)
		Toluene-d8	50	51.5	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.1	106	70 (77)	130 (121)
VX0610WBSD0	VX0610WBSD01	1,2-Dichloroethane-d4	50	49.3	99	70 (74)	130 (125)
		Dibromofluoromethane	50	52.3	105	70 (75)	130 (124)
		Toluene-d8	50	51.0	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.4	105	70 (77)	130 (121)

