

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q2234 SAS No.: Q2234 SDG NO.: Q2234
 Lab File ID: VX046586.D Date Analyzed: 06/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:07
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	66941	12.018				
UPPER LIMIT	133882	12.518				
LOWER LIMIT	33470.5	11.518				
EPA SAMPLE NO.						
MW-17B-55-060425	116641	12.02				
MW-17B-55-060425DL	112148	12.02				
VX0610WBL01	104374	12.02				
VX0610WBS01	68222	12.02				
VX0610WBSD01	65284	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.