

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2235

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136021

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.1	0.099	101	85-115	06/05/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	06/05/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	06/05/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	06/05/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	06/05/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2235

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136032

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	06/05/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	06/05/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	06/05/2025
Sample ID: Corrosivity	CCV3	pH	2.01	2.00	101	90-110	06/05/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2235

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136033

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	06/05/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	06/05/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	06/05/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2235

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136044

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 ASTM Ammonia	mg/L	1	1	100	90-110	06/06/2025
Sample ID: CCV1 ASTM Ammonia	mg/L	1	1	100	90-110	06/06/2025
Sample ID: CCV2 ASTM Ammonia	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV3 ASTM Ammonia	mg/L	1.1	1	110	90-110	06/06/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q2235

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136065

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV ASTM COD	mg/L	50.962	50	102	95-105	05/28/2025
Sample ID: CCV1 ASTM COD	mg/L	47.915	50	96	95-105	06/09/2025
Sample ID: CCV2 ASTM COD	mg/L	50.962	50	102	95-105	06/09/2025
Sample ID: CCV3 ASTM COD	mg/L	49.946	50	100	95-105	06/09/2025