

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2235

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136057

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136057BS							
ASTM Oil and Grease	mg/L	20.0	16.8		84	1	78-114	06/09/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2235

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136065

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136065BS							
ASTM COD	mg/L	50	52.0		104	1	90-110	06/09/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2235

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136073

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136073BS							
Oil and Grease	mg/Kg	100	94.9		95	1	80-120	06/10/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2235

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136044

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB168331BS							
ASTM Ammonia	mg/L	1	1.00		100	1	90-110	06/06/2025