

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q2235
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2235 **SAS No.:** Q2235
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV31	Mercury	4.04	4.0	101	90 - 110	CV	06/10/2025	10:39	LB136079

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q2235
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2235 **SAS No.:** Q2235
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV17	Mercury	4.62	5.0	92	90 - 110	CV	06/10/2025	10:43	LB136079
CCV18	Mercury	4.87	5.0	98	90 - 110	CV	06/10/2025	11:32	LB136079
CCV19	Mercury	5.40	5.0	108	90 - 110	CV	06/10/2025	12:12	LB136079
CCV20	Mercury	4.67	5.0	93	90 - 110	CV	06/10/2025	12:57	LB136079
CCV21	Mercury	5.40	5.0	108	90 - 110	CV	06/10/2025	13:30	LB136079
CCV22	Mercury	4.85	5.0	97	90 - 110	CV	06/10/2025	14:28	LB136079

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q2235
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2235 **SAS No.:** Q2235
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	964	1000	96	90 - 110	P	06/09/2025	12:31	LB136071
	Barium	477	520	92	90 - 110	P	06/09/2025	12:31	LB136071
	Cadmium	472	510	93	90 - 110	P	06/09/2025	12:31	LB136071
	Chromium	504	520	97	90 - 110	P	06/09/2025	12:31	LB136071
	Copper	505	510	99	90 - 110	P	06/09/2025	12:31	LB136071
	Lead	927	1000	93	90 - 110	P	06/09/2025	12:31	LB136071
	Nickel	505	530	95	90 - 110	P	06/09/2025	12:31	LB136071
	Selenium	947	1000	95	90 - 110	P	06/09/2025	12:31	LB136071
	Silver	256	250	102	90 - 110	P	06/09/2025	12:31	LB136071
	Zinc	967	1000	97	90 - 110	P	06/09/2025	12:31	LB136071

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT SDG No.: Q2235
 Contract: ENTA05 Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	18.9	20.0	95	80 - 120	P	06/09/2025	12:36	LB136071
	Barium	93.5	100	94	80 - 120	P	06/09/2025	12:36	LB136071
	Cadmium	5.83	6.0	97	80 - 120	P	06/09/2025	12:36	LB136071
	Chromium	9.51	10.0	95	80 - 120	P	06/09/2025	12:36	LB136071
	Copper	21.2	20.0	106	80 - 120	P	06/09/2025	12:36	LB136071
	Lead	13.3	12.0	111	80 - 120	P	06/09/2025	12:36	LB136071
	Nickel	39.5	40.0	99	80 - 120	P	06/09/2025	12:36	LB136071
	Selenium	17.9	20.0	89	80 - 120	P	06/09/2025	12:36	LB136071
	Silver	9.82	10.0	98	80 - 120	P	06/09/2025	12:36	LB136071
	Zinc	43.3	40.0	108	80 - 120	P	06/09/2025	12:36	LB136071

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q2235
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2235 **SAS No.:** Q2235
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5000	5000	100	90 - 110	P	06/09/2025	13:06	LB136071
	Barium	9520	10000	95	90 - 110	P	06/09/2025	13:06	LB136071
	Cadmium	2440	2500	98	90 - 110	P	06/09/2025	13:06	LB136071
	Chromium	997	1000	100	90 - 110	P	06/09/2025	13:06	LB136071
	Copper	1260	1250	100	90 - 110	P	06/09/2025	13:06	LB136071
	Lead	4900	5000	98	90 - 110	P	06/09/2025	13:06	LB136071
	Nickel	2460	2500	98	90 - 110	P	06/09/2025	13:06	LB136071
	Selenium	5060	5000	101	90 - 110	P	06/09/2025	13:06	LB136071
	Silver	1250	1250	100	90 - 110	P	06/09/2025	13:06	LB136071
	Zinc	2490	2500	100	90 - 110	P	06/09/2025	13:06	LB136071
CCV02	Arsenic	4990	5000	100	90 - 110	P	06/09/2025	13:53	LB136071
	Barium	9530	10000	95	90 - 110	P	06/09/2025	13:53	LB136071
	Cadmium	2440	2500	98	90 - 110	P	06/09/2025	13:53	LB136071
	Chromium	987	1000	99	90 - 110	P	06/09/2025	13:53	LB136071
	Copper	1260	1250	100	90 - 110	P	06/09/2025	13:53	LB136071
	Lead	4900	5000	98	90 - 110	P	06/09/2025	13:53	LB136071
	Nickel	2460	2500	98	90 - 110	P	06/09/2025	13:53	LB136071
	Selenium	5060	5000	101	90 - 110	P	06/09/2025	13:53	LB136071
	Silver	1230	1250	99	90 - 110	P	06/09/2025	13:53	LB136071
	Zinc	2460	2500	98	90 - 110	P	06/09/2025	13:53	LB136071
CCV03	Arsenic	4930	5000	98	90 - 110	P	06/09/2025	14:39	LB136071
	Barium	9250	10000	92	90 - 110	P	06/09/2025	14:39	LB136071
	Cadmium	2390	2500	96	90 - 110	P	06/09/2025	14:39	LB136071
	Chromium	958	1000	96	90 - 110	P	06/09/2025	14:39	LB136071
	Copper	1240	1250	99	90 - 110	P	06/09/2025	14:39	LB136071
	Lead	4800	5000	96	90 - 110	P	06/09/2025	14:39	LB136071
	Nickel	2410	2500	96	90 - 110	P	06/09/2025	14:39	LB136071
	Selenium	5010	5000	100	90 - 110	P	06/09/2025	14:39	LB136071
	Silver	1200	1250	96	90 - 110	P	06/09/2025	14:39	LB136071
	Zinc	2400	2500	96	90 - 110	P	06/09/2025	14:39	LB136071
CCV04	Arsenic	4800	5000	96	90 - 110	P	06/09/2025	15:34	LB136071
	Barium	9450	10000	94	90 - 110	P	06/09/2025	15:34	LB136071
	Cadmium	2330	2500	93	90 - 110	P	06/09/2025	15:34	LB136071
	Chromium	961	1000	96	90 - 110	P	06/09/2025	15:34	LB136071

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT SDG No.: Q2235
Contract: ENTA05 Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Copper	1210	1250	97	90 - 110	P	06/09/2025	15:34	LB136071
	Lead	4700	5000	94	90 - 110	P	06/09/2025	15:34	LB136071
	Nickel	2360	2500	94	90 - 110	P	06/09/2025	15:34	LB136071
	Selenium	4870	5000	97	90 - 110	P	06/09/2025	15:34	LB136071
	Silver	1210	1250	97	90 - 110	P	06/09/2025	15:34	LB136071
	Zinc	2430	2500	97	90 - 110	P	06/09/2025	15:34	LB136071
CCV05	Arsenic	4780	5000	96	90 - 110	P	06/09/2025	16:23	LB136071
	Barium	9510	10000	95	90 - 110	P	06/09/2025	16:23	LB136071
	Cadmium	2350	2500	94	90 - 110	P	06/09/2025	16:23	LB136071
	Chromium	953	1000	95	90 - 110	P	06/09/2025	16:23	LB136071
	Copper	1210	1250	96	90 - 110	P	06/09/2025	16:23	LB136071
	Lead	4730	5000	95	90 - 110	P	06/09/2025	16:23	LB136071
	Nickel	2370	2500	95	90 - 110	P	06/09/2025	16:23	LB136071
	Selenium	4840	5000	97	90 - 110	P	06/09/2025	16:23	LB136071
	Silver	1190	1250	95	90 - 110	P	06/09/2025	16:23	LB136071
	Zinc	2370	2500	95	90 - 110	P	06/09/2025	16:23	LB136071

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT SDG No.: Q2235
 Contract: ENTA05 Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	3850	4000	96	90 - 110	P	06/10/2025	14:37	LB136097
	Barium	7450	8000	93	90 - 110	P	06/10/2025	14:37	LB136097
	Cadmium	1920	2000	96	90 - 110	P	06/10/2025	14:37	LB136097
	Chromium	794	800	99	90 - 110	P	06/10/2025	14:37	LB136097
	Copper	994	1000	99	90 - 110	P	06/10/2025	14:37	LB136097
	Lead	3800	4000	95	90 - 110	P	06/10/2025	14:37	LB136097
	Nickel	1930	2000	97	90 - 110	P	06/10/2025	14:37	LB136097
	Selenium	3880	4000	97	90 - 110	P	06/10/2025	14:37	LB136097
	Silver	978	1000	98	90 - 110	P	06/10/2025	14:37	LB136097
	Zinc	1980	2000	99	90 - 110	P	06/10/2025	14:37	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT SDG No.: Q2235
Contract: ENTA05 Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	21.2	20.0	106	80 - 120	P	06/10/2025	14:41	LB136097
	Barium	101	100	101	80 - 120	P	06/10/2025	14:41	LB136097
	Cadmium	6.31	6.0	105	80 - 120	P	06/10/2025	14:41	LB136097
	Chromium	10.8	10.0	108	80 - 120	P	06/10/2025	14:41	LB136097
	Copper	22.6	20.0	113	80 - 120	P	06/10/2025	14:41	LB136097
	Lead	12.8	12.0	107	80 - 120	P	06/10/2025	14:41	LB136097
	Nickel	41.6	40.0	104	80 - 120	P	06/10/2025	14:41	LB136097
	Selenium	19.7	20.0	98	80 - 120	P	06/10/2025	14:41	LB136097
	Silver	10.5	10.0	105	80 - 120	P	06/10/2025	14:41	LB136097
	Zinc	43.7	40.0	109	80 - 120	P	06/10/2025	14:41	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q2235
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2235 **SAS No.:** Q2235
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5050	5000	101	90 - 110	P	06/10/2025	15:30	LB136097
	Barium	9840	10000	98	90 - 110	P	06/10/2025	15:30	LB136097
	Cadmium	2510	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	15:30	LB136097
	Copper	1270	1250	101	90 - 110	P	06/10/2025	15:30	LB136097
	Lead	5030	5000	100	90 - 110	P	06/10/2025	15:30	LB136097
	Nickel	2510	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Selenium	5080	5000	102	90 - 110	P	06/10/2025	15:30	LB136097
	Silver	1270	1250	101	90 - 110	P	06/10/2025	15:30	LB136097
	Zinc	2560	2500	102	90 - 110	P	06/10/2025	15:30	LB136097
CCV02	Arsenic	5070	5000	101	90 - 110	P	06/10/2025	16:28	LB136097
	Barium	9830	10000	98	90 - 110	P	06/10/2025	16:28	LB136097
	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	16:28	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	16:28	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	16:28	LB136097
	Lead	5070	5000	101	90 - 110	P	06/10/2025	16:28	LB136097
	Nickel	2540	2500	101	90 - 110	P	06/10/2025	16:28	LB136097
	Selenium	5100	5000	102	90 - 110	P	06/10/2025	16:28	LB136097
	Silver	1270	1250	102	90 - 110	P	06/10/2025	16:28	LB136097
	Zinc	2570	2500	103	90 - 110	P	06/10/2025	16:28	LB136097
CCV03	Arsenic	5130	5000	103	90 - 110	P	06/10/2025	17:36	LB136097
	Barium	10200	10000	102	90 - 110	P	06/10/2025	17:36	LB136097
	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	17:36	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	17:36	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	17:36	LB136097
	Lead	5080	5000	102	90 - 110	P	06/10/2025	17:36	LB136097
	Nickel	2540	2500	102	90 - 110	P	06/10/2025	17:36	LB136097
	Selenium	5140	5000	103	90 - 110	P	06/10/2025	17:36	LB136097
	Silver	1280	1250	103	90 - 110	P	06/10/2025	17:36	LB136097
	Zinc	2580	2500	103	90 - 110	P	06/10/2025	17:36	LB136097
CCV04	Arsenic	5200	5000	104	90 - 110	P	06/10/2025	18:28	LB136097
	Barium	10300	10000	103	90 - 110	P	06/10/2025	18:28	LB136097
	Cadmium	2580	2500	103	90 - 110	P	06/10/2025	18:28	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	18:28	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q2235
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2235 **SAS No.:** Q2235
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Copper	1290	1250	103	90 - 110	P	06/10/2025	18:28	LB136097
	Lead	5160	5000	103	90 - 110	P	06/10/2025	18:28	LB136097
	Nickel	2570	2500	103	90 - 110	P	06/10/2025	18:28	LB136097
	Selenium	5220	5000	104	90 - 110	P	06/10/2025	18:28	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	18:28	LB136097
	Zinc	2540	2500	101	90 - 110	P	06/10/2025	18:28	LB136097
CCV05	Arsenic	5170	5000	103	90 - 110	P	06/10/2025	19:18	LB136097
	Barium	10200	10000	102	90 - 110	P	06/10/2025	19:18	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	19:18	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	19:18	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	19:18	LB136097
	Lead	5090	5000	102	90 - 110	P	06/10/2025	19:18	LB136097
	Nickel	2540	2500	102	90 - 110	P	06/10/2025	19:18	LB136097
	Selenium	5190	5000	104	90 - 110	P	06/10/2025	19:18	LB136097
	Silver	1260	1250	101	90 - 110	P	06/10/2025	19:18	LB136097
	Zinc	2510	2500	100	90 - 110	P	06/10/2025	19:18	LB136097
CCV06	Arsenic	5160	5000	103	90 - 110	P	06/10/2025	20:09	LB136097
	Barium	9960	10000	100	90 - 110	P	06/10/2025	20:09	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	20:09	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	20:09	LB136097
	Lead	5110	5000	102	90 - 110	P	06/10/2025	20:09	LB136097
	Nickel	2550	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Selenium	5190	5000	104	90 - 110	P	06/10/2025	20:09	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	20:09	LB136097
	Zinc	2570	2500	103	90 - 110	P	06/10/2025	20:09	LB136097
CCV07	Arsenic	5230	5000	104	90 - 110	P	06/10/2025	21:04	LB136097
	Barium	10100	10000	101	90 - 110	P	06/10/2025	21:04	LB136097
	Cadmium	2600	2500	104	90 - 110	P	06/10/2025	21:04	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	21:04	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	21:04	LB136097
	Lead	5210	5000	104	90 - 110	P	06/10/2025	21:04	LB136097
	Nickel	2590	2500	104	90 - 110	P	06/10/2025	21:04	LB136097
	Selenium	5230	5000	105	90 - 110	P	06/10/2025	21:04	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** Q2235
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2235 **SAS No.:** Q2235
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Silver	1280	1250	102	90 - 110	P	06/10/2025	21:04	LB136097
	Zinc	2540	2500	102	90 - 110	P	06/10/2025	21:04	LB136097
CCV08	Arsenic	5240	5000	105	90 - 110	P	06/10/2025	22:04	LB136097
	Barium	10100	10000	101	90 - 110	P	06/10/2025	22:04	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	22:04	LB136097
	Copper	1280	1250	103	90 - 110	P	06/10/2025	22:04	LB136097
	Lead	5130	5000	102	90 - 110	P	06/10/2025	22:04	LB136097
	Nickel	2560	2500	103	90 - 110	P	06/10/2025	22:04	LB136097
	Selenium	5280	5000	106	90 - 110	P	06/10/2025	22:04	LB136097
	Silver	1280	1250	102	90 - 110	P	06/10/2025	22:04	LB136097
	Zinc	2550	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
CCV09	Arsenic	5150	5000	103	90 - 110	P	06/10/2025	22:56	LB136097
	Barium	10100	10000	101	90 - 110	P	06/10/2025	22:56	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	22:56	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	22:56	LB136097
	Lead	5100	5000	102	90 - 110	P	06/10/2025	22:56	LB136097
	Nickel	2560	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Selenium	5180	5000	104	90 - 110	P	06/10/2025	22:56	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	22:56	LB136097
	Zinc	2580	2500	103	90 - 110	P	06/10/2025	22:56	LB136097
CCV10	Arsenic	5180	5000	104	90 - 110	P	06/10/2025	23:57	LB136097
	Barium	10000	10000	100	90 - 110	P	06/10/2025	23:57	LB136097
	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	23:57	LB136097
	Chromium	1060	1000	106	90 - 110	P	06/10/2025	23:57	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	23:57	LB136097
	Lead	5100	5000	102	90 - 110	P	06/10/2025	23:57	LB136097
	Nickel	2560	2500	102	90 - 110	P	06/10/2025	23:57	LB136097
	Selenium	5230	5000	105	90 - 110	P	06/10/2025	23:57	LB136097
	Silver	1310	1250	104	90 - 110	P	06/10/2025	23:57	LB136097
	Zinc	2630	2500	105	90 - 110	P	06/10/2025	23:57	LB136097
CCV11	Arsenic	5280	5000	106	90 - 110	P	06/11/2025	01:35	LB136097
	Barium	10100	10000	101	90 - 110	P	06/11/2025	01:35	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT SDG No.: Q2235
Contract: ENTA05 Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV11	Cadmium	2570	2500	103	90 - 110	P	06/11/2025	01:35	LB136097
	Chromium	1040	1000	104	90 - 110	P	06/11/2025	01:35	LB136097
	Copper	1290	1250	103	90 - 110	P	06/11/2025	01:35	LB136097
	Lead	5160	5000	103	90 - 110	P	06/11/2025	01:35	LB136097
	Nickel	2590	2500	104	90 - 110	P	06/11/2025	01:35	LB136097
	Selenium	5320	5000	106	90 - 110	P	06/11/2025	01:35	LB136097
	Silver	1280	1250	103	90 - 110	P	06/11/2025	01:35	LB136097
	Zinc	2600	2500	104	90 - 110	P	06/11/2025	01:35	LB136097
CCV12	Arsenic	4850	5000	97	90 - 110	P	06/11/2025	02:21	LB136097
	Barium	10000	10000	100	90 - 110	P	06/11/2025	02:21	LB136097
	Cadmium	2400	2500	96	90 - 110	P	06/11/2025	02:21	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/11/2025	02:21	LB136097
	Copper	1200	1250	96	90 - 110	P	06/11/2025	02:21	LB136097
	Lead	4810	5000	96	90 - 110	P	06/11/2025	02:21	LB136097
	Nickel	2420	2500	97	90 - 110	P	06/11/2025	02:21	LB136097
	Selenium	4880	5000	98	90 - 110	P	06/11/2025	02:21	LB136097
	Silver	1280	1250	102	90 - 110	P	06/11/2025	02:21	LB136097
	Zinc	2590	2500	104	90 - 110	P	06/11/2025	02:21	LB136097