



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT SDG No.: Q2235
Contract: ENTA05 Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	19.0	20.0	95	65 - 135	P	06/09/2025	12:44	LB136071
	Barium	91.7	100	92	65 - 135	P	06/09/2025	12:44	LB136071
	Cadmium	5.74	6.0	96	65 - 135	P	06/09/2025	12:44	LB136071
	Chromium	9.13	10.0	91	65 - 135	P	06/09/2025	12:44	LB136071
	Copper	21.3	20.0	106	65 - 135	P	06/09/2025	12:44	LB136071
	Lead	12.3	12.0	103	65 - 135	P	06/09/2025	12:44	LB136071
	Nickel	38.7	40.0	97	65 - 135	P	06/09/2025	12:44	LB136071
	Selenium	18.5	20.0	93	65 - 135	P	06/09/2025	12:44	LB136071
	Silver	9.51	10.0	95	65 - 135	P	06/09/2025	12:44	LB136071
	Zinc	41.5	40.0	104	65 - 135	P	06/09/2025	12:44	LB136071
CRA	Mercury	0.20	0.2	102	70 - 130	CV	06/10/2025	10:56	LB136079
CRI01	Arsenic	22.9	20.0	114	65 - 135	P	06/10/2025	14:50	LB136097
	Barium	98.7	100	99	65 - 135	P	06/10/2025	14:50	LB136097
	Cadmium	6.07	6.0	101	65 - 135	P	06/10/2025	14:50	LB136097
	Chromium	10.7	10.0	107	65 - 135	P	06/10/2025	14:50	LB136097
	Copper	22.7	20.0	114	65 - 135	P	06/10/2025	14:50	LB136097
	Lead	13.4	12.0	112	65 - 135	P	06/10/2025	14:50	LB136097
	Nickel	40.8	40.0	102	65 - 135	P	06/10/2025	14:50	LB136097
	Selenium	19.5	20.0	98	65 - 135	P	06/10/2025	14:50	LB136097
	Silver	10.6	10.0	106	65 - 135	P	06/10/2025	14:50	LB136097
	Zinc	44.1	40.0	110	65 - 135	P	06/10/2025	14:50	LB136097