

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q2235
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2235 **SAS No.:** Q2235
Matrix: Water **Sample ID:** Q2242-04 **Client ID:** TP09-MHJDUP
Percent Solids for Sample: NA **Duplicate ID** Q2242-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	758		799		5		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Copper	ug/L	20	53.5	J	100	U	200.0		P
Lead	ug/L	20	22.8	J	27.3	J	18		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Nickel	ug/L	20	200	U	200	U			P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P
Zinc	ug/L	20	235		256		9		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q2235
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2235 **SAS No.:** Q2235
Matrix: Water **Sample ID:** Q2242-04MS **Client ID:** TP09-MHJMSD
Percent Solids for Sample: NA **Duplicate ID** Q2242-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4020		4150		3		P
Barium	ug/L	20	2090		2060		1		P
Cadmium	ug/L	20	1020		1040		2		P
Chromium	ug/L	20	1980		2010		2		P
Copper	ug/L	20	1480		1530		3		P
Lead	ug/L	20	4890		5020		3		P
Mercury	ug/L	20	30.2		34.8		14		CV
Nickel	ug/L	20	2530		2600		3		P
Selenium	ug/L	20	9960		10300		3		P
Silver	ug/L	20	380		386		2		P
Zinc	ug/L	20	1390		1400		1		P