

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

Continuing Calib Date: 06/07/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 01:37 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1242-2	(2)	4.79	4.79	4.69	4.89	0.00
Aroclor-1242-3	(3)	4.84	4.85	4.75	4.95	0.01
Aroclor-1242-4	(4)	4.96	4.97	4.87	5.07	0.01
Aroclor-1242-5	(5)	5.62	5.62	5.52	5.72	0.00
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.71	8.73	8.63	8.83	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

Continuing Calib Date: 06/07/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 01:37 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.75	4.76	4.66	4.86	0.01
Aroclor-1242-2	(2)	4.77	4.78	4.68	4.88	0.01
Aroclor-1242-3	(3)	4.95	4.95	4.85	5.05	0.00
Aroclor-1242-4	(4)	5.03	5.04	4.94	5.14	0.01
Aroclor-1242-5	(5)	5.55	5.56	5.46	5.66	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.01
Decachlorobiphenyl		8.67	8.67	8.57	8.77	0.00