

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

Continuing Calib Date: 06/07/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 02:13 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.57       | 5.58      | 5.48      | 5.68 | 0.01       |
| Aroclor-1254-2       | (2) | 5.72       | 5.73      | 5.63      | 5.83 | 0.01       |
| Aroclor-1254-3       | (3) | 6.13       | 6.13      | 6.03      | 6.23 | 0.00       |
| Aroclor-1254-4       | (4) | 6.36       | 6.36      | 6.26      | 6.46 | 0.00       |
| Aroclor-1254-5       | (5) | 6.78       | 6.78      | 6.68      | 6.88 | 0.00       |
| Tetrachloro-m-xylene |     | 3.68       | 3.68      | 3.58      | 3.78 | 0.00       |
| Decachlorobiphenyl   |     | 8.71       | 8.72      | 8.62      | 8.82 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

Continuing Calib Date: 06/07/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 02:13 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.55       | 5.56      | 5.46      | 5.66 | 0.01       |
| Aroclor-1254-2 (2)   | 5.70       | 5.71      | 5.61      | 5.81 | 0.01       |
| Aroclor-1254-3 (3)   | 6.10       | 6.11      | 6.01      | 6.21 | 0.01       |
| Aroclor-1254-4 (4)   | 6.33       | 6.34      | 6.24      | 6.44 | 0.01       |
| Aroclor-1254-5 (5)   | 6.75       | 6.75      | 6.65      | 6.85 | 0.00       |
| Tetrachloro-m-xylene | 3.68       | 3.68      | 3.58      | 3.78 | 0.01       |
| Decachlorobiphenyl   | 8.67       | 8.67      | 8.57      | 8.77 | 0.01       |