

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 06/06/2025

Lab Sample No.: AR1660CCC500 Data File : PO111512.D Time Analyzed: 09:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.768	4.673	4.873	547.840	500.000	9.6
Aroclor-1016-2	4.786	4.692	4.892	556.620	500.000	11.3
Aroclor-1016-3	4.843	4.749	4.949	571.660	500.000	14.3
Aroclor-1016-4	4.963	4.869	5.069	564.700	500.000	12.9
Aroclor-1016-5	5.219	5.126	5.326	559.150	500.000	11.8
Aroclor-1260-1	6.260	6.166	6.366	571.690	500.000	14.3
Aroclor-1260-2	6.449	6.355	6.555	608.120	500.000	21.6
Aroclor-1260-3	6.816	6.722	6.922	597.250	500.000	19.5
Aroclor-1260-4	7.075	6.983	7.183	549.490	500.000	9.9
Aroclor-1260-5	7.318	7.225	7.425	555.300	500.000	11.1
Decachlorobiphenyl	8.714	8.624	8.824	48.360	50.000	-3.3
Tetrachloro-m-xylene	3.679	3.582	3.782	53.810	50.000	7.6

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL01 Date Analyzed: 06/06/2025

Lab Sample No.: AR1660CCC500 Data File : PO111512.D Time Analyzed: 09:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.753	4.658	4.858	502.040	500.000	0.4
Aroclor-1016-2	4.771	4.677	4.877	502.610	500.000	0.5
Aroclor-1016-3	4.946	4.852	5.052	504.210	500.000	0.8
Aroclor-1016-4	4.988	4.894	5.094	495.270	500.000	-0.9
Aroclor-1016-5	5.201	5.107	5.307	505.520	500.000	1.1
Aroclor-1260-1	6.231	6.138	6.338	491.740	500.000	-1.7
Aroclor-1260-2	6.420	6.326	6.526	488.760	500.000	-2.2
Aroclor-1260-3	6.572	6.479	6.679	452.270	500.000	-9.5
Aroclor-1260-4	7.042	6.949	7.149	466.220	500.000	-6.8
Aroclor-1260-5	7.284	7.191	7.391	451.920	500.000	-9.6
Decachlorobiphenyl	8.665	8.574	8.774	44.760	50.000	-10.5
Tetrachloro-m-xylene	3.675	3.579	3.779	52.460	50.000	4.9