

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 06/06/2025

Lab Sample No.: AR1660CCC500 Data File : PO111527.D Time Analyzed: 16:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.768	4.673	4.873	540.270	500.000	8.1
Aroclor-1016-2	4.787	4.692	4.892	538.270	500.000	7.7
Aroclor-1016-3	4.843	4.749	4.949	509.890	500.000	2.0
Aroclor-1016-4	4.963	4.869	5.069	505.830	500.000	1.2
Aroclor-1016-5	5.221	5.126	5.326	534.500	500.000	6.9
Aroclor-1260-1	6.260	6.166	6.366	565.590	500.000	13.1
Aroclor-1260-2	6.449	6.355	6.555	559.540	500.000	11.9
Aroclor-1260-3	6.816	6.722	6.922	518.140	500.000	3.6
Aroclor-1260-4	7.076	6.983	7.183	475.330	500.000	-4.9
Aroclor-1260-5	7.319	7.225	7.425	477.670	500.000	-4.5
Decachlorobiphenyl	8.717	8.624	8.824	40.950	50.000	-18.1
Tetrachloro-m-xylene	3.680	3.582	3.782	52.080	50.000	4.2

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/14/2025 05/14/2025

Client Sample No.: CCAL02 Date Analyzed: 06/06/2025

Lab Sample No.: AR1660CCC500 Data File : PO111527.D Time Analyzed: 16:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.754	4.658	4.858	505.020	500.000	1.0
Aroclor-1016-2	4.773	4.677	4.877	504.590	500.000	0.9
Aroclor-1016-3	4.948	4.852	5.052	498.620	500.000	-0.3
Aroclor-1016-4	4.991	4.894	5.094	463.210	500.000	-7.4
Aroclor-1016-5	5.203	5.107	5.307	500.590	500.000	0.1
Aroclor-1260-1	6.233	6.138	6.338	457.600	500.000	-8.5
Aroclor-1260-2	6.421	6.326	6.526	454.580	500.000	-9.1
Aroclor-1260-3	6.573	6.479	6.679	419.480	500.000	-16.1
Aroclor-1260-4	7.043	6.949	7.149	429.720	500.000	-14.1
Aroclor-1260-5	7.285	7.191	7.391	416.590	500.000	-16.7
Decachlorobiphenyl	8.667	8.574	8.774	42.330	50.000	-15.3
Tetrachloro-m-xylene	3.676	3.579	3.779	52.200	50.000	4.4