

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP072683.D Time Analyzed: 09:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.651	5.552	5.752	429.730	500.000	-14.1
Aroclor-1016-2	5.672	5.574	5.774	469.690	500.000	-6.1
Aroclor-1016-3	5.734	5.636	5.836	468.640	500.000	-6.3
Aroclor-1016-4	5.831	5.734	5.934	490.580	500.000	-1.9
Aroclor-1016-5	6.124	6.027	6.227	479.200	500.000	-4.2
Aroclor-1260-1	7.242	7.147	7.347	465.310	500.000	-6.9
Aroclor-1260-2	7.495	7.401	7.601	459.510	500.000	-8.1
Aroclor-1260-3	7.853	7.760	7.960	481.040	500.000	-3.8
Aroclor-1260-4	8.077	7.985	8.185	472.890	500.000	-5.4
Aroclor-1260-5	8.396	8.303	8.503	489.620	500.000	-2.1
Decachlorobiphenyl	10.197	10.113	10.313	50.880	50.000	1.8
Tetrachloro-m-xylene	4.498	4.396	4.596	48.980	50.000	-2.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/06/2025

Lab Sample No.: AR1660CCC500 Data File : PP072683.D Time Analyzed: 09:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.874	4.777	4.977	487.010	500.000	-2.6
Aroclor-1016-2	4.892	4.795	4.995	500.980	500.000	0.2
Aroclor-1016-3	5.069	4.973	5.173	504.430	500.000	0.9
Aroclor-1016-4	5.111	5.015	5.215	512.830	500.000	2.6
Aroclor-1016-5	5.324	5.229	5.429	506.610	500.000	1.3
Aroclor-1260-1	6.356	6.264	6.464	516.180	500.000	3.2
Aroclor-1260-2	6.545	6.453	6.653	494.350	500.000	-1.1
Aroclor-1260-3	6.697	6.606	6.806	521.800	500.000	4.4
Aroclor-1260-4	7.167	7.077	7.277	532.310	500.000	6.5
Aroclor-1260-5	7.410	7.319	7.519	525.020	500.000	5.0
Decachlorobiphenyl	8.808	8.723	8.923	54.550	50.000	9.1
Tetrachloro-m-xylene	3.793	3.691	3.891	49.280	50.000	-1.4