

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05
Lab Code: CHEM **Case No.:** Q2235 **SAS No.:** Q2235 **SDG NO.:** Q2235
Instrument ID: ECD_S **Calibration Date(s):** 06/04/2025 06/04/2025
Calibration Times: 11:19 12:55

GC Column: RTX-CLP **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 = <u>PS030476.D</u>	CF 500 = <u>PS030477.D</u>				
CF 750 = <u>PS030478.D</u>		CF 1000 = <u>PS030479.D</u>	CF 1500 = <u>PS030480.D</u>				
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	23444500000	20930200000	20252400000	19918900000	18779400000	20665100000	8
2,4-D	4208920000	3650560000	3531620000	3494410000	3332550000	3643610000	9
2,4-DCAA	4505990000	3745160000	3587820000	3627070000	3409690000	3775150000	11

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: ENTA05
Lab Code: CHEM **Case No.:** Q2235 **SAS No.:** Q2235 **SDG NO.:** Q2235
Instrument ID: ECD_S **Calibration Date(s):** 06/04/2025 06/04/2025
Calibration Times: 11:19 12:55

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 = <u>PS030476.D</u>	CF 500 = <u>PS030477.D</u>				
CF 750 = <u>PS030478.D</u>		CF 1000 = <u>PS030479.D</u>	CF 1500 = <u>PS030480.D</u>				
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	15627400000	13837500000	13337200000	13203400000	12578300000	13716700000	8
2,4-D	1858730000	1575260000	1504430000	1488570000	1434610000	1572320000	11
2,4-DCAA	1341960000	1066160000	1011260000	997042000	961634000	1075610000	14