

Surrogate Summary

SDG No.: Q2235

Client: ENTACT

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS030475.D	PIBLK-PS030475.D	2,4-DCAA	1	500	414	83		70 (61)	130 (136)
		2,4-DCAA	2	500	445	89		70 (61)	130 (136)
I.BLK-PS030534.D	PIBLK-PS030534.D	2,4-DCAA	1	500	455	91		70 (61)	130 (136)
		2,4-DCAA	2	500	469	94		70 (61)	130 (136)
PB168329BL	PB168329BL	2,4-DCAA	1	500	425	85		70 (61)	130 (136)
		2,4-DCAA	2	500	484	97		70 (61)	130 (136)
PB168329BS	PB168329BS	2,4-DCAA	1	500	523	105		70 (61)	130 (136)
		2,4-DCAA	2	500	506	101		70 (61)	130 (136)
I.BLK-PS030546.D	PIBLK-PS030546.D	2,4-DCAA	1	500	454	91		70 (61)	130 (136)
		2,4-DCAA	2	500	474	95		70 (61)	130 (136)
PB168311TB	PB168311TB	2,4-DCAA	1	500	437	87		70 (61)	130 (136)
		2,4-DCAA	2	500	433	87		70 (61)	130 (136)
I.BLK-PS030559.D	PIBLK-PS030559.D	2,4-DCAA	1	500	458	92		70 (61)	130 (136)
		2,4-DCAA	2	500	474	95		70 (61)	130 (136)
I.BLK-PS030562.D	PIBLK-PS030562.D	2,4-DCAA	1	500	433	87		70 (61)	130 (136)
		2,4-DCAA	2	500	447	89		70 (61)	130 (136)
Q2235-03	WC-A2-08-C	2,4-DCAA	1	500	400	80		70 (61)	130 (136)
		2,4-DCAA	2	500	349	70		70 (61)	130 (136)
Q2198-04MS	B-207-SB02MS	2,4-DCAA	1	500	614	123		70 (61)	130 (136)
		2,4-DCAA	2	500	531	106		70 (61)	130 (136)
Q2198-04MSD	B-207-SB02MSD	2,4-DCAA	1	500	625	125		70 (61)	130 (136)
		2,4-DCAA	2	500	543	109		70 (61)	130 (136)
I.BLK-PS030574.D	PIBLK-PS030574.D	2,4-DCAA	1	500	437	87		70 (61)	130 (136)
		2,4-DCAA	2	500	454	91		70 (61)	130 (136)