

Surrogate Summary

SDG No.: Q2235

Client: ENTACT

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088583.D	PIBLK-PD088583.D	Decachlorobiphenyl	1	20	18.4	92		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	16.5	83		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	18.0	90		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	16.9	84		30 (61)	150 (148)
I.BLK-PD088820.D	PIBLK-PD088820.D	Decachlorobiphenyl	1	20	24.1	120		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	23.7	119		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	23.0	115		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	21.4	107		30 (61)	150 (148)
PB168330BL	PB168330BL	Decachlorobiphenyl	1	20	25.8	129		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	24.1	121		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	24.4	122		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	21.8	109		30 (61)	150 (148)
PB168330BS	PB168330BS	Decachlorobiphenyl	1	20	23.8	119		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	22.0	110		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	22.9	115		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	20.4	102		30 (61)	150 (148)
PB168271TB	PB168271TB	Decachlorobiphenyl	1	20	25.4	127		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	23.6	118		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	24.2	121		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	21.6	108		30 (61)	150 (148)
Q2198-04MS	B-207-SB02MS	Decachlorobiphenyl	1	20	25.9	129		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	24.2	121		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	24.8	124		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	21.1	105		30 (61)	150 (148)
Q2198-04MSD	B-207-SB02MSD	Decachlorobiphenyl	1	20	26.1	131		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	24.3	122		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	24.9	124		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	21.3	106		30 (61)	150 (148)
I.BLK-PD088840.D	PIBLK-PD088840.D	Decachlorobiphenyl	1	20	24.7	123		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	23.3	117		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	23.7	118		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	20.8	104		30 (61)	150 (148)
I.BLK-PD088853.D	PIBLK-PD088853.D	Decachlorobiphenyl	1	20	25.2	126		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	24.7	123		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	24.6	123		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	22.5	113		30 (61)	150 (148)
Q2235-03	WC-A2-08-C	Decachlorobiphenyl	1	20	23.9	120		30 (57)	150 (171)
		Tetrachloro-m-xylene	1	20	21.5	108		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	21.6	108		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	19.6	98		30 (61)	150 (148)
I.BLK-PD088865.D	PIBLK-PD088865.D	Decachlorobiphenyl	1	20	25.0	125		30 (57)	150 (171)

Surrogate Summary

SDG No.: Q2235

Client: ENTACT

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088865.D	PIBLK-PD088865.D	Tetrachloro-m-xylene	1	20	24.2	121		30 (61)	150 (148)
		Decachlorobiphenyl	2	20	24.3	121		30 (57)	150 (171)
		Tetrachloro-m-xylene	2	20	22.3	112		30 (61)	150 (148)