

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/10/2025

Lab Sample No.: PSTDCCC050 Data File : PD088855.D Time Analyzed: 10:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.082	8.977	9.177	52.360	50.000	4.7
Endrin	6.582	6.477	6.677	52.340	50.000	4.7
gamma-BHC (Lindane)	4.337	4.232	4.432	53.690	50.000	7.4
Heptachlor	4.936	4.832	5.032	54.030	50.000	8.1
Heptachlor epoxide	5.698	5.593	5.793	52.050	50.000	4.1
Methoxychlor	7.501	7.395	7.595	52.680	50.000	5.4
Tetrachloro-m-xylene	3.556	3.451	3.651	57.030	50.000	14.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 06/10/2025

Lab Sample No.: PSTDCCC050 Data File : PD088855.D Time Analyzed: 10:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.075	7.977	8.177	50.310	50.000	0.6
Endrin	5.790	5.692	5.892	47.150	50.000	-5.7
gamma-BHC (Lindane)	3.729	3.631	3.831	48.270	50.000	-3.5
Heptachlor	4.083	3.985	4.185	47.410	50.000	-5.2
Heptachlor epoxide	4.873	4.775	4.975	47.790	50.000	-4.4
Methoxychlor	6.756	6.658	6.858	47.840	50.000	-4.3
Tetrachloro-m-xylene	2.879	2.782	2.982	47.950	50.000	-4.1