

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 06/10/2025

Lab Sample No.: PSTDCCC050 Data File : PD088866.D Time Analyzed: 14:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.073	8.977	9.177	51.320	50.000	2.6
Endrin	6.574	6.477	6.677	51.740	50.000	3.5
gamma-BHC (Lindane)	4.329	4.232	4.432	53.320	50.000	6.6
Heptachlor	4.928	4.832	5.032	53.600	50.000	7.2
Heptachlor epoxide	5.690	5.593	5.793	52.350	50.000	4.7
Methoxychlor	7.493	7.395	7.595	51.440	50.000	2.9
Tetrachloro-m-xylene	3.548	3.451	3.651	57.070	50.000	14.1

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL04 Date Analyzed: 06/10/2025

Lab Sample No.: PSTDCCC050 Data File : PD088866.D Time Analyzed: 14:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.073	7.977	8.177	49.750	50.000	-0.5
Endrin	5.789	5.692	5.892	47.940	50.000	-4.1
gamma-BHC (Lindane)	3.729	3.631	3.831	48.820	50.000	-2.4
Heptachlor	4.083	3.985	4.185	48.230	50.000	-3.5
Heptachlor epoxide	4.872	4.775	4.975	48.250	50.000	-3.5
Methoxychlor	6.755	6.658	6.858	48.550	50.000	-2.9
Tetrachloro-m-xylene	2.880	2.782	2.982	48.570	50.000	-2.9