

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/10/2025

Lab File ID: BM050262.D Time Analyzed: 04:48

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	460715	7.763	1877130	10.55	1070950	14.40
UPPER LIMIT	921430	8.263	3754260	11.051	2141900	14.898
LOWER LIMIT	230358	7.263	938565	10.051	535475	13.898
EPA SAMPLE NO.						
01 PB168352BL	334540	7.76	1240840	10.55	698531	14.39
02 PB168352BS	333914	7.76	1311550	10.55	754658	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.