

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG No.: Q2235

Instrument ID: MSVOA_N Calibration Date/Time: 06/09/2025 08:37

Lab File ID: VN086888.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.664	0.635		-4.37	20
1,1-Dichloroethene	0.557	0.556		-0.18	20
2-Butanone	0.577	0.476		-17.5	20
Carbon Tetrachloride	0.433	0.431		-0.46	20
Chloroform	1.118	1.055		-5.64	20
Benzene	1.444	1.423		-1.45	20
1,2-Dichloroethane	0.438	0.420		-4.11	20
Trichloroethene	0.342	0.352		2.92	20
Tetrachloroethene	0.316	0.318		0.63	20
Chlorobenzene	1.103	1.140	0.3	3.35	20
1,2-Dichloroethane-d4	0.669	0.579		-13.45	20
Dibromofluoromethane	0.296	0.295		-0.34	20
Toluene-d8	1.173	1.125		-4.09	20
4-Bromofluorobenzene	0.436	0.405		-7.11	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.