

Surrogate Summary

SDG No.: Q2235

Client: ENTACT

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2235-01	WC-A2-08-G	1,2-Dichloroethane-d4	50	44.2	88	70 (74)	130 (125)
		Dibromofluoromethane	50	44.9	90	70 (75)	130 (124)
		Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.0	98	70 (77)	130 (121)

Surrogate Summary

SDG No.: Q2235

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN0609WBL01	VN0609WBL01	1,2-Dichloroethane-d4	50	42.8	86	70 (74)	130 (125)
		Dibromofluoromethane	50	48.1	96	70 (75)	130 (124)
		Toluene-d8	50	50.7	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.9	98	70 (77)	130 (121)
VN0609WBS01	VN0609WBS01	1,2-Dichloroethane-d4	50	56.6	113	70 (74)	130 (125)
		Dibromofluoromethane	50	58.6	117	70 (75)	130 (124)
		Toluene-d8	50	55.2	110	70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.7	111	70 (77)	130 (121)
VN0609WBSD0	VN0609WBSD01	1,2-Dichloroethane-d4	50	44.5	89	70 (74)	130 (125)
		Dibromofluoromethane	50	50.5	101	70 (75)	130 (124)
		Toluene-d8	50	48.2	96	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.7	95	70 (77)	130 (121)

