

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235
 Lab File ID: VN086888.D Date Analyzed: 06/09/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:37
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	261450	8.23	451735	9.11	380922	11.87
UPPER LIMIT	522900	8.729	903470	9.606	761844	12.365
LOWER LIMIT	130725	7.729	225868	8.606	190461	11.365
EPA SAMPLE NO.						
WC-A2-08-G	386024	8.24	684260	9.11	589833	11.87
VN0609WBL01	390256	8.23	684606	9.11	591728	11.87
VN0609WBS01	266986	8.23	481053	9.11	422488	11.87
VN0609WBSD01	261604	8.23	462540	9.11	392287	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.