

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q2235 SAS No.: Q2235 SDG NO.: Q2235
 Lab File ID: VN086888.D Date Analyzed: 06/09/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:37
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	178512	13.788				
UPPER LIMIT	357024	14.288				
LOWER LIMIT	89256	13.288				
EPA SAMPLE NO.						
WC-A2-08-G	286619	13.79				
VN0609WBL01	288224	13.79				
VN0609WBS01	202169	13.79				
VN0609WBSD01	182953	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.