

Hit Summary Sheet
SW-846

SDG No.: Q2236
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	WC-A4-05A-G							
Q2236-01	WC-A4-05A-G	TCLP	Vinyl Chloride	0.0031	J	0.00026	0.0050	mg/L
Q2236-01	WC-A4-05A-G	TCLP	Benzene	0.025		0.00015	0.0050	mg/L
			Total Voc :	0.028				
			Total Concentration:	0.028				
Client ID:	WC-A2-04-G							
Q2236-05	WC-A2-04-G	TCLP	Benzene	0.11		0.00015	0.0050	mg/L
			Total Voc :	0.11				
			Total Concentration:	0.11				
Client ID:	WC-A2-04-GRE							
Q2236-05RE	WC-A2-04-GRE	TCLP	Benzene	0.11		0.00015	0.0050	mg/L
			Total Voc :	0.11				
			Total Concentration:	0.11				
Client ID:	WC-A2-05-G							
Q2236-09	WC-A2-05-G	TCLP	Vinyl Chloride	0.0024	J	0.00026	0.0050	mg/L
Q2236-09	WC-A2-05-G	TCLP	Benzene	0.023		0.00015	0.0050	mg/L
			Total Voc :	0.025				
			Total Concentration:	0.025				
Client ID:	WC-A2-07-G							
Q2236-17	WC-A2-07-G	TCLP	Benzene	0.063		0.00015	0.0050	mg/L
			Total Voc :	0.063				
			Total Concentration:	0.063				
Client ID:	WC-A2-07-GRE							
Q2236-17RE	WC-A2-07-GRE	TCLP	Benzene	0.051		0.00015	0.0050	mg/L
			Total Voc :	0.051				
			Total Concentration:	0.051				



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A4-05A-G	SDG No.:	Q2236
Lab Sample ID:	Q2236-01	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086897.D	1		06/09/25 12:16	VN060925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.0031	J	0.00026	0.0050	mg/L
75-35-4	1,1-Dichloroethene	0.00023	U	0.00023	0.0050	mg/L
78-93-3	2-Butanone	0.00098	U	0.00098	0.025	mg/L
56-23-5	Carbon Tetrachloride	0.00025	U	0.00025	0.0050	mg/L
67-66-3	Chloroform	0.00025	U	0.00025	0.0050	mg/L
71-43-2	Benzene	0.025		0.00015	0.0050	mg/L
107-06-2	1,2-Dichloroethane	0.00022	U	0.00022	0.0050	mg/L
79-01-6	Trichloroethene	0.000090	U	0.000090	0.0050	mg/L
127-18-4	Tetrachloroethene	0.00023	U	0.00023	0.0050	mg/L
108-90-7	Chlorobenzene	0.00012	U	0.00012	0.0050	mg/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	43.3		70 (74) - 130 (125)	87%	SPK: 50
1868-53-7	Dibromofluoromethane	42.8		70 (75) - 130 (124)	86%	SPK: 50
2037-26-5	Toluene-d8	51.3		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.9		70 (77) - 130 (121)	98%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	386000	8.23			
540-36-3	1,4-Difluorobenzene	679000	9.106			
3114-55-4	Chlorobenzene-d5	591000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	281000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-04-G	SDG No.:	Q2236
Lab Sample ID:	Q2236-05	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046552.D	1		06/07/25 03:26	VX060625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.00026	U	0.00026	0.0050	mg/L
75-35-4	1,1-Dichloroethene	0.00023	U	0.00023	0.0050	mg/L
78-93-3	2-Butanone	0.00098	U	0.00098	0.025	mg/L
56-23-5	Carbon Tetrachloride	0.00025	U	0.00025	0.0050	mg/L
67-66-3	Chloroform	0.00025	U	0.00025	0.0050	mg/L
71-43-2	Benzene	0.11		0.00015	0.0050	mg/L
107-06-2	1,2-Dichloroethane	0.00022	U	0.00022	0.0050	mg/L
79-01-6	Trichloroethene	0.000090	U	0.000090	0.0050	mg/L
127-18-4	Tetrachloroethene	0.00023	U	0.00023	0.0050	mg/L
108-90-7	Chlorobenzene	0.00012	U	0.00012	0.0050	mg/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.4		70 (74) - 130 (125)	99%	SPK: 50
1868-53-7	Dibromofluoromethane	33.8	*	70 (75) - 130 (124)	68%	SPK: 50
2037-26-5	Toluene-d8	53.0		70 (86) - 130 (113)	106%	SPK: 50
460-00-4	4-Bromofluorobenzene	57.6		70 (77) - 130 (121)	115%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	117000	5.568			
540-36-3	1,4-Difluorobenzene	225000	6.775			
3114-55-4	Chlorobenzene-d5	237000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	118000	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-04-GRE	SDG No.:	Q2236
Lab Sample ID:	Q2236-05RE	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086898.D	1		06/09/25 12:38	VN060925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.00026	U	0.00026	0.0050	mg/L
75-35-4	1,1-Dichloroethene	0.00023	U	0.00023	0.0050	mg/L
78-93-3	2-Butanone	0.00098	U	0.00098	0.025	mg/L
56-23-5	Carbon Tetrachloride	0.00025	U	0.00025	0.0050	mg/L
67-66-3	Chloroform	0.00025	U	0.00025	0.0050	mg/L
71-43-2	Benzene	0.11		0.00015	0.0050	mg/L
107-06-2	1,2-Dichloroethane	0.00022	U	0.00022	0.0050	mg/L
79-01-6	Trichloroethene	0.000090	U	0.000090	0.0050	mg/L
127-18-4	Tetrachloroethene	0.00023	U	0.00023	0.0050	mg/L
108-90-7	Chlorobenzene	0.00012	U	0.00012	0.0050	mg/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	43.2		70 (74) - 130 (125)	86%	SPK: 50
1868-53-7	Dibromofluoromethane	32.6	*	70 (75) - 130 (124)	65%	SPK: 50
2037-26-5	Toluene-d8	51.3		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.6		70 (77) - 130 (121)	97%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	357000	8.229			
540-36-3	1,4-Difluorobenzene	625000	9.106			
3114-55-4	Chlorobenzene-d5	548000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	258000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-05-G	SDG No.:	Q2236
Lab Sample ID:	Q2236-09	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086899.D	1		06/09/25 12:59	VN060925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.0024	J	0.00026	0.0050	mg/L
75-35-4	1,1-Dichloroethene	0.00023	U	0.00023	0.0050	mg/L
78-93-3	2-Butanone	0.00098	U	0.00098	0.025	mg/L
56-23-5	Carbon Tetrachloride	0.00025	U	0.00025	0.0050	mg/L
67-66-3	Chloroform	0.00025	U	0.00025	0.0050	mg/L
71-43-2	Benzene	0.023		0.00015	0.0050	mg/L
107-06-2	1,2-Dichloroethane	0.00022	U	0.00022	0.0050	mg/L
79-01-6	Trichloroethene	0.000090	U	0.000090	0.0050	mg/L
127-18-4	Tetrachloroethene	0.00023	U	0.00023	0.0050	mg/L
108-90-7	Chlorobenzene	0.00012	U	0.00012	0.0050	mg/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	41.5		70 (74) - 130 (125)	83%	SPK: 50
1868-53-7	Dibromofluoromethane	40.2		70 (75) - 130 (124)	80%	SPK: 50
2037-26-5	Toluene-d8	50.6		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.2		70 (77) - 130 (121)	98%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	396000	8.23			
540-36-3	1,4-Difluorobenzene	681000	9.106			
3114-55-4	Chlorobenzene-d5	597000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	291000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-06-G	SDG No.:	Q2236
Lab Sample ID:	Q2236-13	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046554.D	1		06/07/25 04:09	VX060625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.00026	U	0.00026	0.0050	mg/L
75-35-4	1,1-Dichloroethene	0.00023	U	0.00023	0.0050	mg/L
78-93-3	2-Butanone	0.00098	U	0.00098	0.025	mg/L
56-23-5	Carbon Tetrachloride	0.00025	U	0.00025	0.0050	mg/L
67-66-3	Chloroform	0.00025	U	0.00025	0.0050	mg/L
71-43-2	Benzene	0.00015	U	0.00015	0.0050	mg/L
107-06-2	1,2-Dichloroethane	0.00022	U	0.00022	0.0050	mg/L
79-01-6	Trichloroethene	0.000090	U	0.000090	0.0050	mg/L
127-18-4	Tetrachloroethene	0.00023	U	0.00023	0.0050	mg/L
108-90-7	Chlorobenzene	0.00012	U	0.00012	0.0050	mg/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.2		70 (74) - 130 (125)	92%	SPK: 50
1868-53-7	Dibromofluoromethane	41.7		70 (75) - 130 (124)	83%	SPK: 50
2037-26-5	Toluene-d8	52.6		70 (86) - 130 (113)	105%	SPK: 50
460-00-4	4-Bromofluorobenzene	59.0		70 (77) - 130 (121)	118%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	110000	5.568			
540-36-3	1,4-Difluorobenzene	212000	6.775			
3114-55-4	Chlorobenzene-d5	221000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	117000	12.024			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-07-G	SDG No.:	Q2236
Lab Sample ID:	Q2236-17	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046555.D	1		06/07/25 04:31	VX060625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.00026	U	0.00026	0.0050	mg/L
75-35-4	1,1-Dichloroethene	0.00023	U	0.00023	0.0050	mg/L
78-93-3	2-Butanone	0.00098	U	0.00098	0.025	mg/L
56-23-5	Carbon Tetrachloride	0.00025	U	0.00025	0.0050	mg/L
67-66-3	Chloroform	0.00025	U	0.00025	0.0050	mg/L
71-43-2	Benzene	0.063		0.00015	0.0050	mg/L
107-06-2	1,2-Dichloroethane	0.00022	U	0.00022	0.0050	mg/L
79-01-6	Trichloroethene	0.000090	U	0.000090	0.0050	mg/L
127-18-4	Tetrachloroethene	0.00023	U	0.00023	0.0050	mg/L
108-90-7	Chlorobenzene	0.00012	U	0.00012	0.0050	mg/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	47.5		70 (74) - 130 (125)	95%	SPK: 50
1868-53-7	Dibromofluoromethane	34.9		70 (75) - 130 (124)	70%	SPK: 50
2037-26-5	Toluene-d8	52.8		70 (86) - 130 (113)	106%	SPK: 50
460-00-4	4-Bromofluorobenzene	58.1		70 (77) - 130 (121)	116%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	120000	5.568			
540-36-3	1,4-Difluorobenzene	229000	6.775			
3114-55-4	Chlorobenzene-d5	242000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	121000	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-07-GRE	SDG No.:	Q2236
Lab Sample ID:	Q2236-17RE	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :	SW5035		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086900.D	1		06/09/25 13:21	VN060925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.00026	U	0.00026	0.0050	mg/L
75-35-4	1,1-Dichloroethene	0.00023	U	0.00023	0.0050	mg/L
78-93-3	2-Butanone	0.00098	U	0.00098	0.025	mg/L
56-23-5	Carbon Tetrachloride	0.00025	U	0.00025	0.0050	mg/L
67-66-3	Chloroform	0.00025	U	0.00025	0.0050	mg/L
71-43-2	Benzene	0.051		0.00015	0.0050	mg/L
107-06-2	1,2-Dichloroethane	0.00022	U	0.00022	0.0050	mg/L
79-01-6	Trichloroethene	0.000090	U	0.000090	0.0050	mg/L
127-18-4	Tetrachloroethene	0.00023	U	0.00023	0.0050	mg/L
108-90-7	Chlorobenzene	0.00012	U	0.00012	0.0050	mg/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	42.8		70 (74) - 130 (125)	86%	SPK: 50
1868-53-7	Dibromofluoromethane	31.8	*	70 (75) - 130 (124)	64%	SPK: 50
2037-26-5	Toluene-d8	50.6		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.8		70 (77) - 130 (121)	96%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	440000	8.235			
540-36-3	1,4-Difluorobenzene	778000	9.106			
3114-55-4	Chlorobenzene-d5	671000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	315000	13.788			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q2236	OrderDate:	6/5/2025 11:00:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Austin Farmerie	Location:	N31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2236-01	WC-A4-05A-G	TCLP	TCLP VOA	8260D	06/04/25		06/09/25	06/04/25
Q2236-05	WC-A2-04-G	TCLP	TCLP VOA	8260D	06/04/25		06/07/25	06/04/25
Q2236-05RE	WC-A2-04-GRE	TCLP	TCLP VOA	8260D	06/04/25		06/09/25	06/04/25
Q2236-09	WC-A2-05-G	TCLP	TCLP VOA	8260D	06/04/25		06/09/25	06/04/25
Q2236-13	WC-A2-06-G	TCLP	TCLP VOA	8260D	06/04/25		06/07/25	06/04/25
Q2236-17	WC-A2-07-G	TCLP	TCLP VOA	8260D	06/04/25		06/07/25	06/04/25
Q2236-17RE	WC-A2-07-GRE	TCLP	TCLP VOA	8260D	06/04/25		06/09/25	06/04/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

A

B

C

D

Hit Summary Sheet SW-846

SDG No.: Q2236
Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :				0.000				
			Total Svoc :		0.00			
			Total Concentration:		0.00			



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT		Date Collected:	06/09/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/09/25	
Client Sample ID:	PB168311TB		SDG No.:	Q2236	
Lab Sample ID:	PB168311TB		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050246.D	1	06/09/25 10:45	06/09/25 17:43	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	0.013	U	0.013	0.050	mg/L
106-46-7	1,4-Dichlorobenzene	0.0053	U	0.0053	0.050	mg/L
95-48-7	2-Methylphenol	0.011	U	0.011	0.050	mg/L
65794-96-9	3+4-Methylphenols	0.011	U	0.011	0.10	mg/L
67-72-1	Hexachloroethane	0.0065	U	0.0065	0.050	mg/L
98-95-3	Nitrobenzene	0.0076	U	0.0076	0.050	mg/L
87-68-3	Hexachlorobutadiene	0.0054	U	0.0054	0.050	mg/L
88-06-2	2,4,6-Trichlorophenol	0.0051	U	0.0051	0.050	mg/L
95-95-4	2,4,5-Trichlorophenol	0.0062	U	0.0062	0.050	mg/L
121-14-2	2,4-Dinitrotoluene	0.012	U	0.012	0.050	mg/L
118-74-1	Hexachlorobenzene	0.0052	U	0.0052	0.050	mg/L
87-86-5	Pentachlorophenol	0.016	U	0.016	0.10	mg/L
SURROGATES						
367-12-4	2-Fluorophenol	141		15 (23) - 110 (138)	94%	SPK: 150
13127-88-3	Phenol-d6	132		15 (10) - 110 (134)	88%	SPK: 150
4165-60-0	Nitrobenzene-d5	86.1		30 (67) - 130 (132)	86%	SPK: 100
321-60-8	2-Fluorobiphenyl	83.7		30 (52) - 130 (132)	84%	SPK: 100
118-79-6	2,4,6-Tribromophenol	142		15 (44) - 110 (137)	94%	SPK: 150
1718-51-0	Terphenyl-d14	81.7		30 (42) - 130 (152)	82%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	349000	7.769			
1146-65-2	Naphthalene-d8	1310000	10.551			
15067-26-2	Acenaphthene-d10	755000	14.398			
1517-22-2	Phenanthrene-d10	1470000	17.133			
1719-03-5	Chrysene-d12	1520000	21.368			
1520-96-3	Perylene-d12	1650000	24.344			

Report of Analysis

Client:	ENTACT		Date Collected:	06/09/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/09/25	
Client Sample ID:	PB168311TB		SDG No.:	Q2236	
Lab Sample ID:	PB168311TB		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050246.D	1	06/09/25 10:45	06/09/25 17:43	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A4-05A-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-03		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050253.D	1	06/09/25 10:45	06/09/25 22:17	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	0.013	U	0.013	0.050	mg/L
106-46-7	1,4-Dichlorobenzene	0.0053	U	0.0053	0.050	mg/L
95-48-7	2-Methylphenol	0.011	U	0.011	0.050	mg/L
65794-96-9	3+4-Methylphenols	0.011	U	0.011	0.10	mg/L
67-72-1	Hexachloroethane	0.0065	U	0.0065	0.050	mg/L
98-95-3	Nitrobenzene	0.0076	U	0.0076	0.050	mg/L
87-68-3	Hexachlorobutadiene	0.0054	U	0.0054	0.050	mg/L
88-06-2	2,4,6-Trichlorophenol	0.0051	U	0.0051	0.050	mg/L
95-95-4	2,4,5-Trichlorophenol	0.0062	U	0.0062	0.050	mg/L
121-14-2	2,4-Dinitrotoluene	0.012	U	0.012	0.050	mg/L
118-74-1	Hexachlorobenzene	0.0052	U	0.0052	0.050	mg/L
87-86-5	Pentachlorophenol	0.016	U	0.016	0.10	mg/L
SURROGATES						
367-12-4	2-Fluorophenol	129		15 (23) - 110 (138)	86%	SPK: 150
13127-88-3	Phenol-d6	111		15 (10) - 110 (134)	74%	SPK: 150
4165-60-0	Nitrobenzene-d5	90.1		30 (67) - 130 (132)	90%	SPK: 100
321-60-8	2-Fluorobiphenyl	85.0		30 (52) - 130 (132)	85%	SPK: 100
118-79-6	2,4,6-Tribromophenol	150		15 (44) - 110 (137)	100%	SPK: 150
1718-51-0	Terphenyl-d14	85.1		30 (42) - 130 (152)	85%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	339000	7.769			
1146-65-2	Naphthalene-d8	1290000	10.557			
15067-26-2	Acenaphthene-d10	735000	14.398			
1517-22-2	Phenanthrene-d10	1370000	17.133			
1719-03-5	Chrysene-d12	1410000	21.362			
1520-96-3	Perylene-d12	1490000	24.339			

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A4-05A-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-03		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050253.D	1	06/09/25 10:45	06/09/25 22:17	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-04-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-07		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050248.D	1	06/09/25 10:45	06/09/25 19:01	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	0.013	U	0.013	0.050	mg/L
106-46-7	1,4-Dichlorobenzene	0.0053	U	0.0053	0.050	mg/L
95-48-7	2-Methylphenol	0.011	U	0.011	0.050	mg/L
65794-96-9	3+4-Methylphenols	0.011	U	0.011	0.10	mg/L
67-72-1	Hexachloroethane	0.0065	U	0.0065	0.050	mg/L
98-95-3	Nitrobenzene	0.0076	U	0.0076	0.050	mg/L
87-68-3	Hexachlorobutadiene	0.0054	U	0.0054	0.050	mg/L
88-06-2	2,4,6-Trichlorophenol	0.0051	U	0.0051	0.050	mg/L
95-95-4	2,4,5-Trichlorophenol	0.0062	U	0.0062	0.050	mg/L
121-14-2	2,4-Dinitrotoluene	0.012	U	0.012	0.050	mg/L
118-74-1	Hexachlorobenzene	0.0052	U	0.0052	0.050	mg/L
87-86-5	Pentachlorophenol	0.016	U	0.016	0.10	mg/L
SURROGATES						
367-12-4	2-Fluorophenol	137		15 (23) - 110 (138)	92%	SPK: 150
13127-88-3	Phenol-d6	121		15 (10) - 110 (134)	81%	SPK: 150
4165-60-0	Nitrobenzene-d5	89.7		30 (67) - 130 (132)	90%	SPK: 100
321-60-8	2-Fluorobiphenyl	81.9		30 (52) - 130 (132)	82%	SPK: 100
118-79-6	2,4,6-Tribromophenol	149		15 (44) - 110 (137)	99%	SPK: 150
1718-51-0	Terphenyl-d14	83.5		30 (42) - 130 (152)	84%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	334000	7.769			
1146-65-2	Naphthalene-d8	1310000	10.557			
15067-26-2	Acenaphthene-d10	761000	14.398			
1517-22-2	Phenanthrene-d10	1460000	17.133			
1719-03-5	Chrysene-d12	1560000	21.362			
1520-96-3	Perylene-d12	1690000	24.344			

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-04-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-07		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050248.D	1	06/09/25 10:45	06/09/25 19:01	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-05-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-11		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050254.D	1	06/09/25 10:45	06/09/25 22:56	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	0.013	U	0.013	0.050	mg/L
106-46-7	1,4-Dichlorobenzene	0.0053	U	0.0053	0.050	mg/L
95-48-7	2-Methylphenol	0.011	U	0.011	0.050	mg/L
65794-96-9	3+4-Methylphenols	0.011	U	0.011	0.10	mg/L
67-72-1	Hexachloroethane	0.0065	U	0.0065	0.050	mg/L
98-95-3	Nitrobenzene	0.0076	U	0.0076	0.050	mg/L
87-68-3	Hexachlorobutadiene	0.0054	U	0.0054	0.050	mg/L
88-06-2	2,4,6-Trichlorophenol	0.0051	U	0.0051	0.050	mg/L
95-95-4	2,4,5-Trichlorophenol	0.0062	U	0.0062	0.050	mg/L
121-14-2	2,4-Dinitrotoluene	0.012	U	0.012	0.050	mg/L
118-74-1	Hexachlorobenzene	0.0052	U	0.0052	0.050	mg/L
87-86-5	Pentachlorophenol	0.016	U	0.016	0.10	mg/L
SURROGATES						
367-12-4	2-Fluorophenol	138		15 (23) - 110 (138)	92%	SPK: 150
13127-88-3	Phenol-d6	119		15 (10) - 110 (134)	79%	SPK: 150
4165-60-0	Nitrobenzene-d5	92.7		30 (67) - 130 (132)	93%	SPK: 100
321-60-8	2-Fluorobiphenyl	85.6		30 (52) - 130 (132)	86%	SPK: 100
118-79-6	2,4,6-Tribromophenol	143		15 (44) - 110 (137)	95%	SPK: 150
1718-51-0	Terphenyl-d14	91.3		30 (42) - 130 (152)	91%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	373000	7.769			
1146-65-2	Naphthalene-d8	1420000	10.557			
15067-26-2	Acenaphthene-d10	799000	14.398			
1517-22-2	Phenanthrene-d10	1430000	17.133			
1719-03-5	Chrysene-d12	1320000	21.362			
1520-96-3	Perylene-d12	1350000	24.339			

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-05-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-11		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050254.D	1	06/09/25 10:45	06/09/25 22:56	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-06-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-15		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050255.D	1	06/09/25 10:45	06/09/25 23:35	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	0.013	U	0.013	0.050	mg/L
106-46-7	1,4-Dichlorobenzene	0.0053	U	0.0053	0.050	mg/L
95-48-7	2-Methylphenol	0.011	U	0.011	0.050	mg/L
65794-96-9	3+4-Methylphenols	0.011	U	0.011	0.10	mg/L
67-72-1	Hexachloroethane	0.0065	U	0.0065	0.050	mg/L
98-95-3	Nitrobenzene	0.0076	U	0.0076	0.050	mg/L
87-68-3	Hexachlorobutadiene	0.0054	U	0.0054	0.050	mg/L
88-06-2	2,4,6-Trichlorophenol	0.0051	U	0.0051	0.050	mg/L
95-95-4	2,4,5-Trichlorophenol	0.0062	U	0.0062	0.050	mg/L
121-14-2	2,4-Dinitrotoluene	0.012	U	0.012	0.050	mg/L
118-74-1	Hexachlorobenzene	0.0052	U	0.0052	0.050	mg/L
87-86-5	Pentachlorophenol	0.016	U	0.016	0.10	mg/L
SURROGATES						
367-12-4	2-Fluorophenol	146		15 (23) - 110 (138)	97%	SPK: 150
13127-88-3	Phenol-d6	126		15 (10) - 110 (134)	84%	SPK: 150
4165-60-0	Nitrobenzene-d5	97.0		30 (67) - 130 (132)	97%	SPK: 100
321-60-8	2-Fluorobiphenyl	90.3		30 (52) - 130 (132)	90%	SPK: 100
118-79-6	2,4,6-Tribromophenol	152		15 (44) - 110 (137)	101%	SPK: 150
1718-51-0	Terphenyl-d14	93.6		30 (42) - 130 (152)	94%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	314000	7.769			
1146-65-2	Naphthalene-d8	1210000	10.551			
15067-26-2	Acenaphthene-d10	691000	14.398			
1517-22-2	Phenanthrene-d10	1260000	17.133			
1719-03-5	Chrysene-d12	1230000	21.362			
1520-96-3	Perylene-d12	1310000	24.338			

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-06-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-15		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050255.D	1	06/09/25 10:45	06/09/25 23:35	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-07-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-19		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050256.D	1	06/09/25 10:45	06/10/25 00:15	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	0.013	U	0.013	0.050	mg/L
106-46-7	1,4-Dichlorobenzene	0.0053	U	0.0053	0.050	mg/L
95-48-7	2-Methylphenol	0.011	U	0.011	0.050	mg/L
65794-96-9	3+4-Methylphenols	0.011	U	0.011	0.10	mg/L
67-72-1	Hexachloroethane	0.0065	U	0.0065	0.050	mg/L
98-95-3	Nitrobenzene	0.0076	U	0.0076	0.050	mg/L
87-68-3	Hexachlorobutadiene	0.0054	U	0.0054	0.050	mg/L
88-06-2	2,4,6-Trichlorophenol	0.0051	U	0.0051	0.050	mg/L
95-95-4	2,4,5-Trichlorophenol	0.0062	U	0.0062	0.050	mg/L
121-14-2	2,4-Dinitrotoluene	0.012	U	0.012	0.050	mg/L
118-74-1	Hexachlorobenzene	0.0052	U	0.0052	0.050	mg/L
87-86-5	Pentachlorophenol	0.016	U	0.016	0.10	mg/L
SURROGATES						
367-12-4	2-Fluorophenol	142		15 (23) - 110 (138)	95%	SPK: 150
13127-88-3	Phenol-d6	124		15 (10) - 110 (134)	83%	SPK: 150
4165-60-0	Nitrobenzene-d5	94.4		30 (67) - 130 (132)	94%	SPK: 100
321-60-8	2-Fluorobiphenyl	86.5		30 (52) - 130 (132)	87%	SPK: 100
118-79-6	2,4,6-Tribromophenol	148		15 (44) - 110 (137)	99%	SPK: 150
1718-51-0	Terphenyl-d14	93.9		30 (42) - 130 (152)	94%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	372000	7.769			
1146-65-2	Naphthalene-d8	1450000	10.557			
15067-26-2	Acenaphthene-d10	829000	14.398			
1517-22-2	Phenanthrene-d10	1500000	17.133			
1719-03-5	Chrysene-d12	1400000	21.362			
1520-96-3	Perylene-d12	1430000	24.338			

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-07-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-19		Matrix:	TCLP	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	100	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	TCLP BNA	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :	SW3541				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050256.D	1	06/09/25 10:45	06/10/25 00:15	PB168352

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q2236	OrderDate:	6/5/2025 11:00:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Austin Farmerie	Location:	N31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2236-03	WC-A4-05A-C	TCLP	TCLP BNA	8270E	06/04/25	06/09/25	06/09/25	06/04/25
Q2236-07	WC-A2-04-C	TCLP	TCLP BNA	8270E	06/04/25	06/09/25	06/09/25	06/04/25
Q2236-11	WC-A2-05-C	TCLP	TCLP BNA	8270E	06/04/25	06/09/25	06/09/25	06/04/25
Q2236-15	WC-A2-06-C	TCLP	TCLP BNA	8270E	06/04/25	06/09/25	06/09/25	06/04/25
Q2236-19	WC-A2-07-C	TCLP	TCLP BNA	8270E	06/04/25	06/09/25	06/10/25	06/04/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2236

Order ID: Q2236

Client: ENTACT

Project ID: 540 Degraw St, Brooklyn, NY - E9309

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000

A

B

C

D



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT		Date Collected:		
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/06/25	
Client Sample ID:	PB168311TB		SDG No.:	Q2236	
Lab Sample ID:	PB168311TB		Matrix:	TCLP	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088835.D	1	06/06/25 12:32	06/09/25 13:58	PB168330

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.000037	U	0.000037	0.00050	mg/L
76-44-8	Heptachlor	0.000027	U	0.000027	0.00050	mg/L
1024-57-3	Heptachlor epoxide	0.000096	U	0.000096	0.00050	mg/L
72-20-8	Endrin	0.000032	U	0.000032	0.00050	mg/L
72-43-5	Methoxychlor	0.00011	U	0.00011	0.00050	mg/L
8001-35-2	Toxaphene	0.0017	U	0.0017	0.010	mg/L
57-74-9	Chlordane	0.00088	U	0.00088	0.0050	mg/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	25.4		30 (57) - 150 (171)	127%	SPK: 20
877-09-8	Tetrachloro-m-xylene	23.7		30 (61) - 150 (148)	118%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A4-05A-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-03		Matrix:	TCLP	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088860.D	1	06/06/25 12:32	06/10/25 13:27	PB168330

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.000037	U	0.000037	0.00050	mg/L
76-44-8	Heptachlor	0.000027	U	0.000027	0.00050	mg/L
1024-57-3	Heptachlor epoxide	0.000096	U	0.000096	0.00050	mg/L
72-20-8	Endrin	0.000032	U	0.000032	0.00050	mg/L
72-43-5	Methoxychlor	0.00011	U	0.00011	0.00050	mg/L
8001-35-2	Toxaphene	0.0017	U	0.0017	0.010	mg/L
57-74-9	Chlordane	0.00088	U	0.00088	0.0050	mg/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.4		30 (57) - 150 (171)	107%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.1		30 (61) - 150 (148)	96%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-04-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-07		Matrix:	TCLP	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088861.D	1	06/06/25 12:32	06/10/25 13:41	PB168330

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.000037	U	0.000037	0.00050	mg/L
76-44-8	Heptachlor	0.000027	U	0.000027	0.00050	mg/L
1024-57-3	Heptachlor epoxide	0.000096	U	0.000096	0.00050	mg/L
72-20-8	Endrin	0.000032	U	0.000032	0.00050	mg/L
72-43-5	Methoxychlor	0.00011	U	0.00011	0.00050	mg/L
8001-35-2	Toxaphene	0.0017	U	0.0017	0.010	mg/L
57-74-9	Chlordane	0.00088	U	0.00088	0.0050	mg/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	23.9		30 (57) - 150 (171)	120%	SPK: 20
877-09-8	Tetrachloro-m-xylene	23.1		30 (61) - 150 (148)	116%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-05-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-11		Matrix:	TCLP	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088862.D	1	06/06/25 12:32	06/10/25 13:54	PB168330

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.000037	U	0.000037	0.00050	mg/L
76-44-8	Heptachlor	0.000027	U	0.000027	0.00050	mg/L
1024-57-3	Heptachlor epoxide	0.000096	U	0.000096	0.00050	mg/L
72-20-8	Endrin	0.000032	U	0.000032	0.00050	mg/L
72-43-5	Methoxychlor	0.00011	U	0.00011	0.00050	mg/L
8001-35-2	Toxaphene	0.0017	U	0.0017	0.010	mg/L
57-74-9	Chlordane	0.00088	U	0.00088	0.0050	mg/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	24.7		30 (57) - 150 (171)	124%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.3		30 (61) - 150 (148)	111%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-06-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-15		Matrix:	TCLP	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088863.D	1	06/06/25 12:32	06/10/25 14:08	PB168330

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.000037	U	0.000037	0.00050	mg/L
76-44-8	Heptachlor	0.000027	U	0.000027	0.00050	mg/L
1024-57-3	Heptachlor epoxide	0.000096	U	0.000096	0.00050	mg/L
72-20-8	Endrin	0.000032	U	0.000032	0.00050	mg/L
72-43-5	Methoxychlor	0.00011	U	0.00011	0.00050	mg/L
8001-35-2	Toxaphene	0.0017	U	0.0017	0.010	mg/L
57-74-9	Chlordane	0.00088	U	0.00088	0.0050	mg/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.5		30 (57) - 150 (171)	107%	SPK: 20
877-09-8	Tetrachloro-m-xylene	21.2		30 (61) - 150 (148)	106%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-07-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-19		Matrix:	TCLP	
Analytical Method:	8081B		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PD088864.D	1	06/06/25 12:32	06/10/25 14:22	PB168330

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	0.000037	U	0.000037	0.00050	mg/L
76-44-8	Heptachlor	0.000027	U	0.000027	0.00050	mg/L
1024-57-3	Heptachlor epoxide	0.000096	U	0.000096	0.00050	mg/L
72-20-8	Endrin	0.000032	U	0.000032	0.00050	mg/L
72-43-5	Methoxychlor	0.00011	U	0.00011	0.00050	mg/L
8001-35-2	Toxaphene	0.0017	U	0.0017	0.010	mg/L
57-74-9	Chlordane	0.00088	U	0.00088	0.0050	mg/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	22.9		30 (57) - 150 (171)	115%	SPK: 20
877-09-8	Tetrachloro-m-xylene	22.7		30 (61) - 150 (148)	114%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q2236	OrderDate:	6/5/2025 11:00:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Austin Farmerie	Location:	N31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2236-02	WC-A4-05A-C	SOIL			06/04/25			06/04/25
			PCB	8082A		06/06/25	06/06/25	
Q2236-03	WC-A4-05A-C	TCLP			06/04/25			06/04/25
			TCLP Herbicide	8151A		06/06/25	06/09/25	
			TCLP Pesticide	8081B		06/06/25	06/10/25	
Q2236-06	WC-A2-04-C	SOIL			06/04/25			06/04/25
			PCB	8082A		06/06/25	06/06/25	
Q2236-07	WC-A2-04-C	TCLP			06/04/25			06/04/25
			TCLP Herbicide	8151A		06/06/25	06/09/25	
			TCLP Pesticide	8081B		06/06/25	06/10/25	
Q2236-10	WC-A2-05-C	SOIL			06/04/25			06/04/25
			PCB	8082A		06/06/25	06/06/25	
Q2236-11	WC-A2-05-C	TCLP			06/04/25			06/04/25
			TCLP Herbicide	8151A		06/06/25	06/09/25	
			TCLP Pesticide	8081B		06/06/25	06/10/25	
Q2236-14	WC-A2-06-C	SOIL			06/04/25			06/04/25
			PCB	8082A		06/06/25	06/06/25	
Q2236-15	WC-A2-06-C	TCLP			06/04/25			06/04/25
			TCLP Herbicide	8151A		06/06/25	06/09/25	
			TCLP Pesticide	8081B		06/06/25	06/10/25	
Q2236-18	WC-A2-07-C	SOIL			06/04/25			06/04/25
			PCB	8082A		06/06/25	06/06/25	
Q2236-19	WC-A2-07-C	TCLP			06/04/25			06/04/25
			TCLP Herbicide	8151A		06/06/25	06/09/25	

LAB CHRONICLE

TCLP Pesticide	8081B	06/06/25	06/10/25
----------------	-------	----------	----------

A

C

D



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2236

Order ID: Q2236

Client: ENTACT

Project ID: 540 Degraw St, Brooklyn, NY - E9309

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000

A

B

C

D



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A4-05A-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-02		Matrix:	SOIL	
Analytical Method:	8082A		% Solid:	77.8	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111519.D	1	06/06/25 08:30	06/06/25 13:16	PB168319

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	0.0051	U	0.0051	0.022	mg/Kg
11104-28-2	Aroclor-1221	0.0052	U	0.0052	0.022	mg/Kg
11141-16-5	Aroclor-1232	0.0048	U	0.0048	0.022	mg/Kg
53469-21-9	Aroclor-1242	0.0051	U	0.0051	0.022	mg/Kg
12672-29-6	Aroclor-1248	0.0076	U	0.0076	0.022	mg/Kg
11097-69-1	Aroclor-1254	0.0041	U	0.0041	0.022	mg/Kg
37324-23-5	Aroclor-1262	0.0064	U	0.0064	0.022	mg/Kg
11100-14-4	Aroclor-1268	0.0046	U	0.0046	0.022	mg/Kg
11096-82-5	Aroclor-1260	0.0041	U	0.0041	0.022	mg/Kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.3		30 (32) - 150 (144)	97%	SPK: 20
2051-24-3	Decachlorobiphenyl	12.3		30 (32) - 150 (175)	62%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-04-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-06		Matrix:	SOIL	
Analytical Method:	8082A		% Solid:	71.7	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111520.D	1	06/06/25 08:30	06/06/25 13:35	PB168319

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	0.0055	U	0.0055	0.024	mg/Kg
11104-28-2	Aroclor-1221	0.0056	U	0.0056	0.024	mg/Kg
11141-16-5	Aroclor-1232	0.0052	U	0.0052	0.024	mg/Kg
53469-21-9	Aroclor-1242	0.0056	U	0.0056	0.024	mg/Kg
12672-29-6	Aroclor-1248	0.0082	U	0.0082	0.024	mg/Kg
11097-69-1	Aroclor-1254	0.0045	U	0.0045	0.024	mg/Kg
37324-23-5	Aroclor-1262	0.0070	U	0.0070	0.024	mg/Kg
11100-14-4	Aroclor-1268	0.0050	U	0.0050	0.024	mg/Kg
11096-82-5	Aroclor-1260	0.0045	U	0.0045	0.024	mg/Kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.5		30 (32) - 150 (144)	77%	SPK: 20
2051-24-3	Decachlorobiphenyl	13.9		30 (32) - 150 (175)	70%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-05-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-10		Matrix:	SOIL	
Analytical Method:	8082A		% Solid:	79.2	Decanted:
Sample Wt/Vol:	30.07	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111521.D	1	06/06/25 08:30	06/06/25 13:53	PB168319

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	0.0050	U	0.0050	0.021	mg/Kg
11104-28-2	Aroclor-1221	0.0051	U	0.0051	0.021	mg/Kg
11141-16-5	Aroclor-1232	0.0047	U	0.0047	0.021	mg/Kg
53469-21-9	Aroclor-1242	0.0051	U	0.0051	0.021	mg/Kg
12672-29-6	Aroclor-1248	0.0075	U	0.0075	0.021	mg/Kg
11097-69-1	Aroclor-1254	0.0040	U	0.0040	0.021	mg/Kg
37324-23-5	Aroclor-1262	0.0063	U	0.0063	0.021	mg/Kg
11100-14-4	Aroclor-1268	0.0045	U	0.0045	0.021	mg/Kg
11096-82-5	Aroclor-1260	0.0041	U	0.0041	0.021	mg/Kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.0		30 (32) - 150 (144)	75%	SPK: 20
2051-24-3	Decachlorobiphenyl	9.72		30 (32) - 150 (175)	49%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-06-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-14		Matrix:	SOIL	
Analytical Method:	8082A		% Solid:	83.1	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111522.D	1	06/06/25 08:30	06/06/25 14:12	PB168319

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	0.0047	U	0.0047	0.020	mg/Kg
11104-28-2	Aroclor-1221	0.0048	U	0.0048	0.020	mg/Kg
11141-16-5	Aroclor-1232	0.0045	U	0.0045	0.020	mg/Kg
53469-21-9	Aroclor-1242	0.0048	U	0.0048	0.020	mg/Kg
12672-29-6	Aroclor-1248	0.0071	U	0.0071	0.020	mg/Kg
11097-69-1	Aroclor-1254	0.0039	U	0.0039	0.020	mg/Kg
37324-23-5	Aroclor-1262	0.0060	U	0.0060	0.020	mg/Kg
11100-14-4	Aroclor-1268	0.0043	U	0.0043	0.020	mg/Kg
11096-82-5	Aroclor-1260	0.0039	U	0.0039	0.020	mg/Kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.2		30 (32) - 150 (144)	106%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.4		30 (32) - 150 (175)	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-07-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-18		Matrix:	SOIL	
Analytical Method:	8082A		% Solid:	72.4	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111523.D	1	06/06/25 08:30	06/06/25 14:30	PB168319

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	0.0055	U	0.0055	0.024	mg/Kg
11104-28-2	Aroclor-1221	0.0056	U	0.0056	0.024	mg/Kg
11141-16-5	Aroclor-1232	0.0051	U	0.0051	0.024	mg/Kg
53469-21-9	Aroclor-1242	0.0055	U	0.0055	0.024	mg/Kg
12672-29-6	Aroclor-1248	0.0082	U	0.0082	0.024	mg/Kg
11097-69-1	Aroclor-1254	0.0044	U	0.0044	0.024	mg/Kg
37324-23-5	Aroclor-1262	0.0069	U	0.0069	0.024	mg/Kg
11100-14-4	Aroclor-1268	0.0050	U	0.0050	0.024	mg/Kg
11096-82-5	Aroclor-1260	0.0045	U	0.0045	0.024	mg/Kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.2		30 (32) - 150 (144)	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.5		30 (32) - 150 (175)	83%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q2236	OrderDate:	6/5/2025 11:00:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Austin Farmerie	Location:	N31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2236-02	WC-A4-05A-C	SOIL	PCB	8082A	06/04/25	06/06/25	06/06/25	06/04/25
Q2236-06	WC-A2-04-C	SOIL	PCB	8082A	06/04/25	06/06/25	06/06/25	06/04/25
Q2236-10	WC-A2-05-C	SOIL	PCB	8082A	06/04/25	06/06/25	06/06/25	06/04/25
Q2236-14	WC-A2-06-C	SOIL	PCB	8082A	06/04/25	06/06/25	06/06/25	06/04/25
Q2236-18	WC-A2-07-C	SOIL	PCB	8082A	06/04/25	06/06/25	06/06/25	06/04/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2236

Order ID: Q2236

Client: ENTACT

Project ID: 540 Degraw St, Brooklyn, NY - E9309

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000

A

B

C

D



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT		Date Collected:		
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/06/25	
Client Sample ID:	PB168311TB		SDG No.:	Q2236	
Lab Sample ID:	PB168311TB		Matrix:	TCLP	
Analytical Method:	8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	8151A				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030558.D	1	06/06/25 11:45	06/07/25 03:38	PB168329

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	0.0092	U	0.0092	0.020	mg/L
93-72-1	2,4,5-TP (Silvex)	0.0078	U	0.0078	0.020	mg/L
SURROGATES						
19719-28-9	2,4-DCAA	437		70 (61) - 130 (136)	87%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A4-05A-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-03		Matrix:	TCLP	
Analytical Method:	8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	8151A				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030565.D	1	06/06/25 11:45	06/09/25 13:45	PB168329

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	0.0092	U	0.0092	0.020	mg/L
93-72-1	2,4,5-TP (Silvex)	0.0078	U	0.0078	0.020	mg/L
SURROGATES						
19719-28-9	2,4-DCAA	372		70 (61) - 130 (136)	74%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-04-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-07		Matrix:	TCLP	
Analytical Method:	8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	8151A				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030566.D	1	06/06/25 11:45	06/09/25 14:09	PB168329

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	0.0092	U	0.0092	0.020	mg/L
93-72-1	2,4,5-TP (Silvex)	0.0078	U	0.0078	0.020	mg/L
SURROGATES						
19719-28-9	2,4-DCAA	392		70 (61) - 130 (136)	78%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-05-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-11		Matrix:	TCLP	
Analytical Method:	8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	8151A				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030567.D	1	06/06/25 11:45	06/09/25 14:33	PB168329

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	0.0092	U	0.0092	0.020	mg/L
93-72-1	2,4,5-TP (Silvex)	0.0078	U	0.0078	0.020	mg/L
SURROGATES						
19719-28-9	2,4-DCAA	344	*	70 (61) - 130 (136)	69%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-06-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-15		Matrix:	TCLP	
Analytical Method:	8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	8151A				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030568.D	1	06/06/25 11:45	06/09/25 14:57	PB168329

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	0.0092	U	0.0092	0.020	mg/L
93-72-1	2,4,5-TP (Silvex)	0.0078	U	0.0078	0.020	mg/L
SURROGATES						
19719-28-9	2,4-DCAA	356		70 (61) - 130 (136)	71%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	WC-A2-07-C		SDG No.:	Q2236	
Lab Sample ID:	Q2236-19		Matrix:	TCLP	
Analytical Method:	8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	8151A				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030569.D	1	06/06/25 11:45	06/09/25 15:22	PB168329

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
94-75-7	2,4-D	0.0092	U	0.0092	0.020	mg/L
93-72-1	2,4,5-TP (Silvex)	0.0078	U	0.0078	0.020	mg/L
SURROGATES						
19719-28-9	2,4-DCAA	372		70 (61) - 130 (136)	74%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q2236	OrderDate:	6/5/2025 11:00:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Austin Farmerie	Location:	N31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2236-02	WC-A4-05A-C	SOIL			06/04/25			06/04/25
			PCB	8082A		06/06/25	06/06/25	
Q2236-03	WC-A4-05A-C	TCLP			06/04/25			06/04/25
			TCLP Herbicide	8151A		06/06/25	06/09/25	
			TCLP Pesticide	8081B		06/06/25	06/10/25	
Q2236-06	WC-A2-04-C	SOIL			06/04/25			06/04/25
			PCB	8082A		06/06/25	06/06/25	
Q2236-07	WC-A2-04-C	TCLP			06/04/25			06/04/25
			TCLP Herbicide	8151A		06/06/25	06/09/25	
			TCLP Pesticide	8081B		06/06/25	06/10/25	
Q2236-10	WC-A2-05-C	SOIL			06/04/25			06/04/25
			PCB	8082A		06/06/25	06/06/25	
Q2236-11	WC-A2-05-C	TCLP			06/04/25			06/04/25
			TCLP Herbicide	8151A		06/06/25	06/09/25	
			TCLP Pesticide	8081B		06/06/25	06/10/25	
Q2236-14	WC-A2-06-C	SOIL			06/04/25			06/04/25
			PCB	8082A		06/06/25	06/06/25	
Q2236-15	WC-A2-06-C	TCLP			06/04/25			06/04/25
			TCLP Herbicide	8151A		06/06/25	06/09/25	
			TCLP Pesticide	8081B		06/06/25	06/10/25	
Q2236-18	WC-A2-07-C	SOIL			06/04/25			06/04/25
			PCB	8082A		06/06/25	06/06/25	
Q2236-19	WC-A2-07-C	TCLP			06/04/25			06/04/25
			TCLP Herbicide	8151A		06/06/25	06/09/25	

LAB CHRONICLE

TCLP Pesticide	8081B	06/06/25	06/10/25
----------------	-------	----------	----------

A

C

D

Hit Summary Sheet
SW-846

SDG No.:	Q2236	Order ID:	Q2236
Client:	ENTACT	Project ID:	540 Degraw St, Brooklyn, NY - E9309

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : WC-A4-05A-C								
Q2236-03	WC-A4-05A-C	TCLP	Barium	0.083	J	0.073	0.50	mg/L
Client ID : WC-A2-04-C								
Q2236-07	WC-A2-04-C	TCLP	Barium	0.97		0.073	0.50	mg/L
Q2236-07	WC-A2-04-C	TCLP	Chromium	0.017	J	0.011	0.050	mg/L
Q2236-07	WC-A2-04-C	TCLP	Lead	0.012	J	0.012	0.060	mg/L
Client ID : WC-A2-05-C								
Q2236-11	WC-A2-05-C	TCLP	Barium	0.12	J	0.073	0.50	mg/L
Client ID : WC-A2-06-C								
Q2236-15	WC-A2-06-C	TCLP	Barium	0.24	J	0.073	0.50	mg/L
Q2236-15	WC-A2-06-C	TCLP	Chromium	0.027	J	0.011	0.050	mg/L
Q2236-15	WC-A2-06-C	TCLP	Copper	0.028	J	0.023	0.10	mg/L
Client ID : WC-A2-07-C								
Q2236-19	WC-A2-07-C	TCLP	Barium	0.83		0.073	0.50	mg/L
Q2236-19	WC-A2-07-C	TCLP	Chromium	0.014	J	0.011	0.050	mg/L



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A4-05A-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-03	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-38-2	Arsenic	0.026	U	1	0.026	0.10	mg/L	06/06/25 12:30	06/09/25 15:05	6010D	SW3050
7440-39-3	Barium	0.083	JN	1	0.073	0.50	mg/L	06/06/25 12:30	06/09/25 15:05	6010D	SW3050
7440-43-9	Cadmium	0.0025	U	1	0.0025	0.030	mg/L	06/06/25 12:30	06/09/25 15:05	6010D	SW3050
7440-47-3	Chromium	0.011	U	1	0.011	0.050	mg/L	06/06/25 12:30	06/09/25 15:05	6010D	SW3050
7440-50-8	Copper	0.023	U	1	0.023	0.10	mg/L	06/06/25 12:30	06/09/25 15:05	6010D	SW3050
7439-92-1	Lead	0.012	U	1	0.012	0.060	mg/L	06/06/25 12:30	06/09/25 15:05	6010D	SW3050
7439-97-6	Mercury	0.00076	U	1	0.00076	0.0020	mg/L	06/09/25 15:15	06/10/25 11:29	7470A	
7440-02-0	Nickel	0.015	U	1	0.015	0.20	mg/L	06/06/25 12:30	06/09/25 15:05	6010D	SW3050
7782-49-2	Selenium	0.048	U	1	0.048	0.10	mg/L	06/06/25 12:30	06/09/25 15:05	6010D	SW3050
7440-22-4	Silver	0.0081	U	1	0.0081	0.050	mg/L	06/06/25 12:30	06/09/25 15:05	6010D	SW3050
7440-66-6	Zinc	0.083	U	1	0.083	0.20	mg/L	06/06/25 12:30	06/09/25 15:05	6010D	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP-FULL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-04-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-07	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-38-2	Arsenic	0.026	U	1	0.026	0.10	mg/L	06/06/25 12:30	06/09/25 15:10	6010D	SW3050
7440-39-3	Barium	0.97	N	1	0.073	0.50	mg/L	06/06/25 12:30	06/09/25 15:10	6010D	SW3050
7440-43-9	Cadmium	0.0025	U	1	0.0025	0.030	mg/L	06/06/25 12:30	06/09/25 15:10	6010D	SW3050
7440-47-3	Chromium	0.017	J	1	0.011	0.050	mg/L	06/06/25 12:30	06/09/25 15:10	6010D	SW3050
7440-50-8	Copper	0.023	U	1	0.023	0.10	mg/L	06/06/25 12:30	06/09/25 15:10	6010D	SW3050
7439-92-1	Lead	0.012	J	1	0.012	0.060	mg/L	06/06/25 12:30	06/09/25 15:10	6010D	SW3050
7439-97-6	Mercury	0.00076	U	1	0.00076	0.0020	mg/L	06/09/25 15:15	06/10/25 11:39	7470A	
7440-02-0	Nickel	0.015	U	1	0.015	0.20	mg/L	06/06/25 12:30	06/09/25 15:10	6010D	SW3050
7782-49-2	Selenium	0.048	U	1	0.048	0.10	mg/L	06/06/25 12:30	06/09/25 15:10	6010D	SW3050
7440-22-4	Silver	0.0081	U	1	0.0081	0.050	mg/L	06/06/25 12:30	06/09/25 15:10	6010D	SW3050
7440-66-6	Zinc	0.083	U	1	0.083	0.20	mg/L	06/06/25 12:30	06/09/25 15:10	6010D	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP-FULL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-05-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-11	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-38-2	Arsenic	0.026	U	1	0.026	0.10	mg/L	06/06/25 12:30	06/09/25 15:15	6010D	SW3050
7440-39-3	Barium	0.12	JN	1	0.073	0.50	mg/L	06/06/25 12:30	06/09/25 15:15	6010D	SW3050
7440-43-9	Cadmium	0.0025	U	1	0.0025	0.030	mg/L	06/06/25 12:30	06/09/25 15:15	6010D	SW3050
7440-47-3	Chromium	0.011	U	1	0.011	0.050	mg/L	06/06/25 12:30	06/09/25 15:15	6010D	SW3050
7440-50-8	Copper	0.023	U	1	0.023	0.10	mg/L	06/06/25 12:30	06/09/25 15:15	6010D	SW3050
7439-92-1	Lead	0.012	U	1	0.012	0.060	mg/L	06/06/25 12:30	06/09/25 15:15	6010D	SW3050
7439-97-6	Mercury	0.00076	U	1	0.00076	0.0020	mg/L	06/09/25 15:15	06/10/25 11:41	7470A	
7440-02-0	Nickel	0.015	U	1	0.015	0.20	mg/L	06/06/25 12:30	06/09/25 15:15	6010D	SW3050
7782-49-2	Selenium	0.048	U	1	0.048	0.10	mg/L	06/06/25 12:30	06/09/25 15:15	6010D	SW3050
7440-22-4	Silver	0.0081	U	1	0.0081	0.050	mg/L	06/06/25 12:30	06/09/25 15:15	6010D	SW3050
7440-66-6	Zinc	0.083	U	1	0.083	0.20	mg/L	06/06/25 12:30	06/09/25 15:15	6010D	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP-FULL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-06-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-15	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-38-2	Arsenic	0.026	U	1	0.026	0.10	mg/L	06/06/25 12:30	06/09/25 15:20	6010D	SW3050
7440-39-3	Barium	0.24	JN	1	0.073	0.50	mg/L	06/06/25 12:30	06/09/25 15:20	6010D	SW3050
7440-43-9	Cadmium	0.0025	U	1	0.0025	0.030	mg/L	06/06/25 12:30	06/09/25 15:20	6010D	SW3050
7440-47-3	Chromium	0.027	J	1	0.011	0.050	mg/L	06/06/25 12:30	06/09/25 15:20	6010D	SW3050
7440-50-8	Copper	0.028	J	1	0.023	0.10	mg/L	06/06/25 12:30	06/09/25 15:20	6010D	SW3050
7439-92-1	Lead	0.012	U	1	0.012	0.060	mg/L	06/06/25 12:30	06/09/25 15:20	6010D	SW3050
7439-97-6	Mercury	0.00076	U	1	0.00076	0.0020	mg/L	06/09/25 15:15	06/10/25 11:43	7470A	
7440-02-0	Nickel	0.015	U	1	0.015	0.20	mg/L	06/06/25 12:30	06/09/25 15:20	6010D	SW3050
7782-49-2	Selenium	0.048	U	1	0.048	0.10	mg/L	06/06/25 12:30	06/09/25 15:20	6010D	SW3050
7440-22-4	Silver	0.0081	U	1	0.0081	0.050	mg/L	06/06/25 12:30	06/09/25 15:20	6010D	SW3050
7440-66-6	Zinc	0.083	U	1	0.083	0.20	mg/L	06/06/25 12:30	06/09/25 15:20	6010D	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP-FULL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-07-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-19	Matrix:	TCLP
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-38-2	Arsenic	0.026	U	1	0.026	0.10	mg/L	06/06/25 12:30	06/09/25 15:24	6010D	SW3050
7440-39-3	Barium	0.83	N	1	0.073	0.50	mg/L	06/06/25 12:30	06/09/25 15:24	6010D	SW3050
7440-43-9	Cadmium	0.0025	U	1	0.0025	0.030	mg/L	06/06/25 12:30	06/09/25 15:24	6010D	SW3050
7440-47-3	Chromium	0.014	J	1	0.011	0.050	mg/L	06/06/25 12:30	06/09/25 15:24	6010D	SW3050
7440-50-8	Copper	0.023	U	1	0.023	0.10	mg/L	06/06/25 12:30	06/09/25 15:24	6010D	SW3050
7439-92-1	Lead	0.012	U	1	0.012	0.060	mg/L	06/06/25 12:30	06/09/25 15:24	6010D	SW3050
7439-97-6	Mercury	0.00076	U	1	0.00076	0.0020	mg/L	06/09/25 15:15	06/10/25 11:46	7470A	
7440-02-0	Nickel	0.015	U	1	0.015	0.20	mg/L	06/06/25 12:30	06/09/25 15:24	6010D	SW3050
7782-49-2	Selenium	0.048	U	1	0.048	0.10	mg/L	06/06/25 12:30	06/09/25 15:24	6010D	SW3050
7440-22-4	Silver	0.0081	U	1	0.0081	0.050	mg/L	06/06/25 12:30	06/09/25 15:24	6010D	SW3050
7440-66-6	Zinc	0.083	U	1	0.083	0.20	mg/L	06/06/25 12:30	06/09/25 15:24	6010D	SW3050

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	TCLP-FULL			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q2236	OrderDate:	6/5/2025 11:00:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Austin Farmerie	Location:	N31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2236-03	WC-A4-05A-C	TCLP			06/04/25			06/04/25
			TCLP Mercury	7470A		06/09/25	06/10/25	
			TCLPMetals Group2	6010D		06/06/25	06/09/25	
Q2236-07	WC-A2-04-C	TCLP			06/04/25			06/04/25
			TCLP Mercury	7470A		06/09/25	06/10/25	
			TCLPMetals Group2	6010D		06/06/25	06/09/25	
Q2236-11	WC-A2-05-C	TCLP			06/04/25			06/04/25
			TCLP Mercury	7470A		06/09/25	06/10/25	
			TCLPMetals Group2	6010D		06/06/25	06/09/25	
Q2236-15	WC-A2-06-C	TCLP			06/04/25			06/04/25
			TCLP Mercury	7470A		06/09/25	06/10/25	
			TCLPMetals Group2	6010D		06/06/25	06/09/25	
Q2236-19	WC-A2-07-C	TCLP			06/04/25			06/04/25
			TCLP Mercury	7470A		06/09/25	06/10/25	
			TCLPMetals Group2	6010D		06/06/25	06/09/25	



SAMPLE DATA

A

B

C

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A4-05A-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-02	Matrix:	SOIL
		% Solid:	77.8

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.
Oil and Grease	405		1	7.46	32.1	mg/Kg		06/10/25 09:15	SW9071B
Paint Filter	1.00	U	1	1.00	1.00	ml/100gm		06/05/25 12:25	9095B
pH	12.1	H	1	0	0	pH		06/05/25 14:47	9045D
TS	77.3		1	1.00	5.00	%		06/05/25 11:00	SM 2540 B-15
TVS	15.5		1	1.00	10.0	%		06/05/25 17:00	160.4

Comments: pH result reported at temperature 26.8 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A4-05A-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-03	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	12.1	H	1	0	0	pH		06/05/25 14:47	9045D
Ignitability	NO		1	0	0	oC		06/05/25 11:17	1030
Reactive Cyanide	0.0084	U	1	0.0084	0.050	mg/Kg	06/05/25 11:45	06/05/25 14:07	9012B
Reactive Sulfide	6.31	J	1	0.20	10.0	mg/Kg	06/05/25 13:45	06/05/25 16:06	9034

Comments: pH result reported at temperature 26.8 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A4-05A-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-04	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
ASTM Ammonia	0.65		1	0.030	0.10	mg/L	06/06/25 12:45	06/06/25 16:04	SM 4500-NH3 B plus NH3 G-11
ASTM COD	43.9		1	1.50	10.0	mg/L		06/09/25 13:49	SM 5220 D-11
ASTM Oil and Grease	0.30	J	1	0.29	5.00	mg/L		06/09/25 12:30	SW1664A
ASTM TS	863		1	1.00	5.00	mg/L		06/06/25 12:00	SM 2540 B-15

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-04-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-06	Matrix:	SOIL
		% Solid:	71.7

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.
Oil and Grease	425		1	8.09	34.8	mg/Kg		06/10/25 09:15	SW9071B
Paint Filter	1.00	U	1	1.00	1.00	ml/100gm		06/05/25 12:32	9095B
pH	12.4	H	1	0	0	pH		06/05/25 14:55	9045D
TS	71.8		1	1.00	5.00	%		06/05/25 11:00	SM 2540 B-15
TVS	21.1		1	1.00	10.0	%		06/05/25 17:00	160.4

Comments: pH result reported at temperature 27.1 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-04-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-07	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	12.4	H	1	0	0	pH		06/05/25 14:55	9045D
Ignitability	NO		1	0	0	oC		06/05/25 11:25	1030
Reactive Cyanide	0.0083	U	1	0.0083	0.049	mg/Kg	06/05/25 11:45	06/05/25 14:07	9012B
Reactive Sulfide	6.39	J	1	0.20	10.0	mg/Kg	06/05/25 13:45	06/05/25 16:08	9034

Comments: pH result reported at temperature 27.1 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-04-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-08	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
ASTM Ammonia	1.20		1	0.030	0.10	mg/L	06/06/25 12:45	06/06/25 16:04	SM 4500-NH3 B plus NH3 G-11
ASTM COD	47.9		1	1.50	10.0	mg/L		06/09/25 13:50	SM 5220 D-11
ASTM Oil and Grease	0.50	J	1	0.29	5.00	mg/L		06/09/25 12:30	SW1664A
ASTM TS	1630		1	1.00	5.00	mg/L		06/06/25 12:00	SM 2540 B-15

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-05-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-10	Matrix:	SOIL
		% Solid:	79.2

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.
Oil and Grease	202		1	7.32	31.5	mg/Kg		06/10/25 09:15	SW9071B
Paint Filter	1.00	U	1	1.00	1.00	ml/100gm		06/05/25 12:40	9095B
pH	12.3	H	1	0	0	pH		06/05/25 15:00	9045D
TS	78.9		1	1.00	5.00	%		06/05/25 11:00	SM 2540 B-15
TVS	13.5		1	1.00	10.0	%		06/05/25 17:00	160.4

Comments: pH result reported at temperature 27.2 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-05-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-11	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	12.3	H	1	0	0	pH		06/05/25 15:00	9045D
Ignitability	NO		1	0	0	oC		06/05/25 11:32	1030
Reactive Cyanide	0.0083	U	1	0.0083	0.050	mg/Kg	06/05/25 11:45	06/05/25 14:07	9012B
Reactive Sulfide	4.73	J	1	0.20	10.0	mg/Kg	06/05/25 13:45	06/05/25 16:10	9034

Comments: pH result reported at temperature 27.2 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-05-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-12	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
ASTM Ammonia	1.00		1	0.030	0.10	mg/L	06/06/25 12:45	06/06/25 16:04	SM 4500-NH3 B plus NH3 G-11
ASTM COD	47.9		1	1.50	10.0	mg/L		06/09/25 13:50	SM 5220 D-11
ASTM Oil and Grease	0.50	J	1	0.29	5.00	mg/L		06/09/25 12:30	SW1664A
ASTM TS	1040		1	1.00	5.00	mg/L		06/06/25 12:00	SM 2540 B-15

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-06-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-14	Matrix:	SOIL
		% Solid:	83.1

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.
Oil and Grease	264		1	6.98	30.0	mg/Kg		06/10/25 09:15	SW9071B
Paint Filter	1.00	U	1	1.00	1.00	ml/100gm		06/05/25 12:48	9095B
pH	12.0	H	1	0	0	pH		06/05/25 15:10	9045D
TS	83.6		1	1.00	5.00	%		06/05/25 11:00	SM 2540 B-15
TVS	5.00	J	1	1.00	10.0	%		06/05/25 17:00	160.4

Comments: pH result reported at temperature 26.6 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-06-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-15	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	12.0	H	1	0	0	pH		06/05/25 15:10	9045D
Ignitability	NO		1	0	0	oC		06/05/25 11:40	1030
Reactive Cyanide	0.0083	U	1	0.0083	0.050	mg/Kg	06/05/25 11:45	06/05/25 14:07	9012B
Reactive Sulfide	4.75	J	1	0.20	10.0	mg/Kg	06/05/25 13:45	06/05/25 16:13	9034

Comments: pH result reported at temperature 26.6 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-06-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-16	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
ASTM Ammonia	0.91		1	0.030	0.10	mg/L	06/06/25 12:45	06/06/25 16:05	SM 4500-NH3 B plus NH3 G-11
ASTM COD	36.7		1	1.50	10.0	mg/L		06/09/25 13:51	SM 5220 D-11
ASTM Oil and Grease	0.29	U	1	0.29	5.00	mg/L		06/09/25 12:30	SW1664A
ASTM TS	11100		1	1.00	5.00	mg/L		06/06/25 12:00	SM 2540 B-15

Comments:

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements
H = Sample Analysis Out Of Hold Time

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-07-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-18	Matrix:	SOIL
		% Solid:	72.4

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.
Oil and Grease	262		1	8.01	34.5	mg/Kg		06/10/25 09:15	SW9071B
Paint Filter	1.00	U	1	1.00	1.00	ml/100gm		06/05/25 12:55	9095B
pH	12.2	H	1	0	0	pH		06/05/25 15:15	9045D
TS	72.7		1	1.00	5.00	%		06/05/25 11:00	SM 2540 B-15
TVS	14.1		1	1.00	10.0	%		06/05/25 17:00	160.4

Comments: pH result reported at temperature 26.8 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-07-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-19	Matrix:	SOIL
		% Solid:	100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	12.2	H	1	0	0	pH		06/05/25 15:15	9045D
Ignitability	NO		1	0	0	oC		06/05/25 11:47	1030
Reactive Cyanide	0.0084	U	1	0.0084	0.050	mg/Kg	06/05/25 11:45	06/05/25 14:07	9012B
Reactive Sulfide	3.16	J	1	0.20	10.0	mg/Kg	06/05/25 13:45	06/05/25 16:15	9034

Comments: pH result reported at temperature 26.8 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	ENTACT	Date Collected:	06/04/25 12:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	06/04/25
Client Sample ID:	WC-A2-07-C	SDG No.:	Q2236
Lab Sample ID:	Q2236-20	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
ASTM Ammonia	0.45		1	0.030	0.10	mg/L	06/06/25 12:45	06/06/25 16:10	SM 4500-NH3 B plus NH3 G-11
ASTM COD	57.1		1	1.50	10.0	mg/L		06/09/25 13:52	SM 5220 D-11
ASTM Oil and Grease	0.30	J	1	0.29	5.00	mg/L		06/09/25 12:30	SW1664A
ASTM TS	1910		1	1.00	5.00	mg/L		06/06/25 12:00	SM 2540 B-15

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

LAB CHRONICLE

OrderID:	Q2236	OrderDate:	6/5/2025 11:00:00 AM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Austin Farmerie	Location:	N31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2236-02	WC-A4-05A-C	SOIL			06/04/25 12:00			06/04/25
			Oil and Grease	9071B			06/10/25 09:15	
			Paint Filter	9095B			06/05/25 12:25	
			pH	9045D			06/05/25 14:47	
			TS	SM2540 B			06/05/25 11:00	
			TVS	160.4			06/05/25 17:00	
Q2236-03	WC-A4-05A-C	SOIL			06/04/25 12:00			06/04/25
			Corrosivity	9045D			06/05/25 14:47	
			Ignitability	1030			06/05/25 11:17	
			Reactive Cyanide	9012B		06/05/25	06/05/25 14:07	
			Reactive Sulfide	9034		06/05/25	06/05/25 16:06	
Q2236-04	WC-A4-05A-C	WATER			06/04/25 12:00			06/04/25
			ASTM Ammonia	SM4500-NH3		06/06/25	06/06/25 16:04	
			ASTM COD	SM5220 D			06/09/25 13:49	
			ASTM Oil and Grease	1664A			06/09/25 12:30	

LAB CHRONICLE

Q2236-06	WC-A2-04-C	SOIL	ASTM TS	SM2540 B	06/04/25 12:00	06/06/25 12:00	06/04/25	
			Oil and Grease	9071B				06/10/25 09:15
			Paint Filter	9095B				06/05/25 12:32
			pH	9045D				06/05/25 14:55
			TS	SM2540 B				06/05/25 11:00
			TVS	160.4				06/05/25 17:00
			Q2236-07	WC-A2-04-C				SOIL
Corrosivity	9045D	06/05/25 11:25						
Ignitability	1030	06/05/25 14:07						
Reactive Cyanide	9012B	06/05/25 16:08						
Reactive Sulfide	9034	06/05/25 16:08						
Q2236-08	WC-A2-04-C	WATER			06/04/25 12:00	06/06/25 16:04	06/04/25	
			ASTM Ammonia	SM4500-NH3				06/06/25 13:50
			ASTM COD	SM5220 D				06/09/25 12:30
			ASTM Oil and Grease	1664A				06/09/25 12:00
			ASTM TS	SM2540 B				06/06/25 12:00
Q2236-10	WC-A2-05-C	SOIL			06/04/25 12:00	06/10/25 09:15	06/04/25	
			Oil and Grease	9071B				

LAB CHRONICLE

Q2236-11	WC-A2-05-C	SOIL	Paint Filter	9095B		06/05/25 12:40
			pH	9045D		06/05/25 15:00
			TS	SM2540 B		06/05/25 11:00
			TVS	160.4		06/05/25 17:00
					06/04/25 12:00	06/04/25
Q2236-12	WC-A2-05-C	WATER	Corrosivity	9045D		06/05/25 15:00
			Ignitability	1030		06/05/25 11:32
			Reactive Cyanide	9012B	06/05/25	06/05/25 14:07
			Reactive Sulfide	9034	06/05/25	06/05/25 16:10
					06/04/25 12:00	06/04/25
Q2236-14	WC-A2-06-C	SOIL	ASTM Ammonia	SM4500-NH3	06/06/25	06/06/25 16:04
			ASTM COD	SM5220 D		06/09/25 13:50
			ASTM Oil and Grease	1664A		06/09/25 12:30
			ASTM TS	SM2540 B		06/06/25 12:00
					06/04/25 12:00	06/04/25
			Oil and Grease	9071B		06/10/25 09:15
			Paint Filter	9095B		06/05/25 12:48
			pH	9045D		06/05/25 15:10
			TS	SM2540 B		06/05/25 11:00

LAB CHRONICLE

Q2236-15	WC-A2-06-C	SOIL	TVS	160.4		06/05/25 17:00	06/04/25 12:00	06/04/25
			Corrosivity	9045D		06/05/25 15:10		
			Ignitability	1030		06/05/25 11:40		
			Reactive Cyanide	9012B	06/05/25	06/05/25 14:07		
			Reactive Sulfide	9034	06/05/25	06/05/25 16:13		
Q2236-16	WC-A2-06-C	WATER	ASTM Ammonia	SM4500-NH3	06/06/25	06/06/25 16:05	06/04/25 12:00	06/04/25
			ASTM COD	SM5220 D		06/09/25 13:51		
			ASTM Oil and Grease	1664A		06/09/25 12:30		
			ASTM TS	SM2540 B		06/06/25 12:00		
Q2236-18	WC-A2-07-C	SOIL	Oil and Grease	9071B		06/10/25 09:15	06/04/25 12:00	06/04/25
			Paint Filter	9095B		06/05/25 12:55		
			pH	9045D		06/05/25 15:15		
			TS	SM2540 B		06/05/25 11:00		
			TVS	160.4		06/05/25 17:00		
Q2236-19	WC-A2-07-C	SOIL					06/04/25 12:00	06/04/25
			Corrosivity	9045D		06/05/25 15:15		

LAB CHRONICLE

Q2236-20	WC-A2-07-C	WATER	Ignitability	1030			06/05/25 11:47	
			Reactive Cyanide	9012B		06/05/25	06/05/25 14:07	
			Reactive Sulfide	9034		06/05/25	06/05/25 16:15	
						06/04/25 12:00		06/04/25
			ASTM Ammonia	SM4500-NH3		06/06/25	06/06/25 16:10	
			ASTM COD	SM5220 D			06/09/25 13:52	
			ASTM Oil and Grease	1664A			06/09/25 12:30	
			ASTM TS	SM2540 B			06/06/25 12:00	