

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2236

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136021

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	06/05/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	06/05/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	06/05/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	06/05/2025
Sample ID: CCB4 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	06/05/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2236

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136044

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	06/06/2025
Sample ID: CCB1 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	06/06/2025
Sample ID: CCB2 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	06/06/2025
Sample ID: CCB3 ASTM Ammonia	mg/L	< 0.0500	0.0500	U	0.030	0.1	06/06/2025

Initial and Continuing Calibration Blank Summary

Client: ENTACT

SDG No.: Q2236

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB136065

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	05/28/2025
Sample ID: CCB1 ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	06/09/2025
Sample ID: CCB2 ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	06/09/2025
Sample ID: CCB3 ASTM COD	mg/L	< 5.0000	5.0000	U	1.50	10	06/09/2025