

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q2236
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2235-04
Client ID:	WC-A2-08-CMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
ASTM Ammonia	mg/L	75-125	1.60		0.52		1	1	108		06/06/2025
ASTM COD	mg/L	75-125	99.7		45.9		50.0	1	108		06/09/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q2236
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2235-04
Client ID:	WC-A2-08-CMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
ASTM Ammonia	mg/L	75-125	1.60		0.52		1	1	108		06/06/2025
ASTM COD	mg/L	75-125	101		45.9		50.0	1	110		06/09/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q2236
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2236-18
Client ID:	WC-A2-07-CMS	Percent Solids for Spike Sample:	72.4

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/Kg	75-125	159		262		138	1	-75	*	06/10/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q2236
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2236-18
Client ID:	WC-A2-07-CMSD	Percent Solids for Spike Sample:	72.4

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/Kg	75-125	145		262		138	1	-85	*	06/10/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q2236
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2236-20
Client ID:	WC-A2-07-CMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
ASTM Oil and Grease	mg/L	78-114	20.3		0.30	J	20.0	1	100		06/09/2025

Matrix Spike Summary

Client:	ENTACT	SDG No.:	Q2236
Project:	540 Degraw St, Brooklyn, NY - E9309	Sample ID:	Q2236-20
Client ID:	WC-A2-07-CMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
ASTM Oil and Grease	mg/L	78-114	20.4		0.30	J	20.0	1	101		06/09/2025