


LB136079

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 3.0117e-004

C= -5.3949e-002

Rho= 0.9997364

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/11
0.0	0.000	-0.052	-0.052	6	0.000	6	0				
0.2	0.200	0.187	-0.013	800	0.0 %	800					-7
2.5	2.500	2.605	0.105	8828	0.0 %	8828					4
5.0	5.000	5.057	0.057	16970	0.0 %	16970					1
7.5	7.500	7.349	-0.151	24579	0.0 %	24579					2
10.0	10.000	10.055	0.055	33565	0.0 %	33565					+1

LB136079 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Jun 2025 09:55:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	6	-	0.0000	7470A	PPB	10 Jun 2025 10:08:33	S	Std
0.2 - 1	502	800	-	0.2000	7470A	PPB	10 Jun 2025 10:10:50	S	Std
2.5 - 1	52.75	8828	-	2.5000	7470A	PPB	10 Jun 2025 10:13:07	S	Std
5.0 - 1	55	16970	-	5.0000	7470A	PPB	10 Jun 2025 10:15:24	S	Std
7.5 - 1	57.75	24579	-	7.5000	7470A	PPB	10 Jun 2025 10:20:30	S	Std
10.0 - 1	57	33565	-	10.0000	7470A	PPB	10 Jun 2025 10:30:58	S	Std
ICV31 - 1	ICV31	13599	4.0417	-	7470A	PPB	10 Jun 2025 10:39:04	U	SMPL
ICB31 - 1	ICB31	-87	-0.0802	-	7470A	PPB	10 Jun 2025 10:41:19	U	SMPL
CCV17 - 1	CCV17	15512	4.6178	-	7470A	PPB	10 Jun 2025 10:43:36	U	SMPL
CCB17 - 1	CCB17	-16	-0.0588	-	7470A	PPB	10 Jun 2025 10:48:34	U	SMPL
CRA - 1	CRA	855	0.2036	-	7470A	PPB	10 Jun 2025 10:56:00	U	SMPL
HighStd - 1	HighStd	36091	10.8156	-	7470A	PPB	10 Jun 2025 11:04:13	U	SMPL
ChkStd - 1	ChkStd	25951	7.7618	-	7470A	PPB	10 Jun 2025 11:06:27	U	SMPL
PB168335BL - 1	PBW	101	-0.0235	-	7470A	PPB	10 Jun 2025 11:11:24	U	SMPL
PB168335BS - 1	LCSW	12170	3.6113	-	7470A	PPB	10 Jun 2025 11:18:31	U	SMPL
Q2226-04 - 1	TP06-MHI-WC	-193	-0.1121	-	7470A	PPB	10 Jun 2025 11:20:47	U	SMPL
Q2227-04 - 1	TP07-MHH-WC	8	-0.0515	-	7470A	PPB	10 Jun 2025 11:23:06	U	SMPL
Q2228-04 - 1	TP08-MHI-WC	103	-0.0229	-	7470A	PPB	10 Jun 2025 11:25:23	U	SMPL
Q2235-03 - 1	WC-A2-08-C	140	-0.0118	-	7470A	PPB	10 Jun 2025 11:27:40	U	SMPL
Q2236-03 - 1	WC-A4-05A-C	125	-0.0163	-	7470A	PPB	10 Jun 2025 11:29:55	U	SMPL
CCV18 - 1	CCV18	16363	4.8741	-	7470A	PPB	10 Jun 2025 11:32:11	U	SMPL
CCB18 - 1	CCB18	32	-0.0443	-	7470A	PPB	10 Jun 2025 11:37:08	U	SMPL
Q2236-07 - 1	WC-A2-04-C	-62	-0.0726	-	7470A	PPB	10 Jun 2025 11:39:23	U	SMPL
Q2236-11 - 1	WC-A2-05-C	-37	-0.0651	-	7470A	PPB	10 Jun 2025 11:41:38	U	SMPL
Q2236-15 - 1	WC-A2-06-C	-74	-0.0762	-	7470A	PPB	10 Jun 2025 11:43:54	U	SMPL
Q2236-19 - 1	WC-A2-07-C	-14	-0.0582	-	7470A	PPB	10 Jun 2025 11:46:09	U	SMPL
Q2240-04 - 1	TP-3	187	0.0024	-	7470A	PPB	10 Jun 2025 11:48:25	U	SMPL
Q2240-08 - 1	TP-2	317	0.0415	-	7470A	PPB	10 Jun 2025 11:50:41	U	SMPL
Q2240-12 - 1	TP-1	212	0.0099	-	7470A	PPB	10 Jun 2025 11:52:58	U	SMPL
Q2241-04 - 1	TP-N	-3	-0.0549	-	7470A	PPB	10 Jun 2025 11:55:15	U	SMPL
Q2241-08 - 1	TP-S	-40	-0.0660	-	7470A	PPB	10 Jun 2025 11:57:33	U	SMPL
Q2242-04 - 1	TP09-MHJ	-31	-0.0633	-	7470A	PPB	10 Jun 2025 11:59:48	U	SMPL
CCV19 - 1	CCV19	18106	5.3991	-	7470A	PPB	10 Jun 2025 12:12:34	U	SMPL
CCB19 - 1	CCB19	-51	-0.0693	-	7470A	PPB	10 Jun 2025 12:14:49	U	SMPL
Q2242-04DUP - 1	TP09-MHJDUP	129	-0.0151	-	7470A	PPB	10 Jun 2025 12:17:07	U	SMPL
Q2242-04MS - 1	TP09-MHJMS	10193	3.0159	-	7470A	PPB	10 Jun 2025 12:30:42	U	SMPL
Q2242-04MSD - 1	TP09-MHJMDS	11738	3.4812	-	7470A	PPB	10 Jun 2025 12:32:56	U	SMPL
PB168364BL - 1	PBW	14	-0.0497	-	7470A	PPB	10 Jun 2025 12:38:35	U	SMPL
PB168364BS - 1	LCSW	13523	4.0188	-	7470A	PPB	10 Jun 2025 12:40:51	U	SMPL
Q2244-04 - 1	TP03-MHC	-51	-0.0693	-	7470A	PPB	10 Jun 2025 12:43:07	U	SMPL
Q2260-04 - 1	TP10-MHG-WC	147	-0.0097	-	7470A	PPB	10 Jun 2025 12:45:24	U	SMPL
Q2262-02 - 1	ARS20-0032	64	-0.0347	-	7470A	PPB	10 Jun 2025 12:47:41	U	SMPL
Q2262-04 - 1	ARS20-0001	-15	-0.0585	-	7470A	PPB	10 Jun 2025 12:49:58	U	SMPL
Q2265-04 - 1	TP-MHL-WC	-63	-0.0729	-	7470A	PPB	10 Jun 2025 12:52:13	U	SMPL
CCV20 - 1	CCV20	15677	4.6675	-	7470A	PPB	10 Jun 2025 12:57:14	U	SMPL
CCB20 - 1	CCB20	186	0.0021	-	7470A	PPB	10 Jun 2025 12:59:29	U	SMPL
Q2266-04 - 1	WC-3	-77	-0.0771	-	7470A	PPB	10 Jun 2025 13:01:47	U	SMPL
Q2266-08 - 1	WC-5	-1	-0.0543	-	7470A	PPB	10 Jun 2025 13:04:03	U	SMPL
Q2266-08DUP - 1	WC-5DUP	-7	-0.0561	-	7470A	PPB	10 Jun 2025 13:06:18	U	SMPL
Q2266-08MS - 1	WC-5MS	10915	3.2333	-	7470A	PPB	10 Jun 2025 13:08:34	U	SMPL
Q2266-08MSD - 1	WC-5MSD	10603	3.1394	-	7470A	PPB	10 Jun 2025 13:10:50	U	SMPL
PB168379BL - 1	PBW	163	-0.0049	-	7470A	PPB	10 Jun 2025 13:13:08	U	SMPL
PB168379BS - 1	LCSW	14359	4.2706	-	7470A	PPB	10 Jun 2025 13:18:41	U	SMPL
Q2237-02 - 1	TW-WTS-10	57	-0.0368	-	7470A	PPB	10 Jun 2025 13:20:58	U	SMPL
Q2237-02DUP - 1	TW-WTS-10DUP	128	-0.0154	-	7470A	PPB	10 Jun 2025 13:23:16	U	SMPL
Q2237-02MS - 1	TW-WTS-10MS	16811	5.0090	-	7470A	PPB	10 Jun 2025 13:25:31	U	SMPL
CCV21 - 1	CCV21	18118	5.4027	-	7470A	PPB	10 Jun 2025 13:30:34	U	SMPL
CCB21 - 1	CCB21	125	-0.0163	-	7470A	PPB	10 Jun 2025 13:32:51	U	SMPL
Q2237-02MSD - 1	TW-WTS-10MSD	16803	5.0066	-	7470A	PPB	10 Jun 2025 13:35:09	U	SMPL
Q2250-05 - 1	EB02-060525	-188	-0.1106	-	7470A	PPB	10 Jun 2025 13:37:24	U	SMPL
PB168311TB - 1		98	-0.0244	-	7470A	PPB	10 Jun 2025 13:39:41	U	SMPL
PB168333TB - 1		123	-0.0169	-	7470A	PPB	10 Jun 2025 13:41:56	U	SMPL
Q2242-04LX5 - 1		39	-0.0422	-	7470A	PPB	10 Jun 2025 13:44:12	U	SMPL
Q2242-04A - 1		15884	4.7299	-	7470A	PPB	10 Jun 2025 13:46:29	U	SMPL
Q2266-08LX5 - 1		-172	-0.1058	-	7470A	PPB	10 Jun 2025 13:48:45	U	SMPL
Q2266-08A - 1		15337	4.5651	-	7470A	PPB	10 Jun 2025 13:58:37	U	SMPL
Q2237-02LX5 - 1		-134	-0.0943	-	7470A	PPB	10 Jun 2025 14:00:54	U	SMPL
Q2237-02A - 1		13743	4.0851	-	7470A	PPB	10 Jun 2025 14:05:32	U	SMPL
CCV22 - 1	CCV22	16297	4.8542	-	7470A	PPB	10 Jun 2025 14:28:21	U	SMPL
CCB22 - 1	CCB22	242	0.0189	-	7470A	PPB	10 Jun 2025 14:30:38	U	SMPL