

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q2236
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2236 **SAS No.:** Q2236

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168309BS							
Arsenic	ug/L	4000	4410		110	80 - 120	P
Barium	ug/L	1000	1100		110	80 - 120	P
Cadmium	ug/L	1000	1110		111	80 - 120	P
Chromium	ug/L	2000	2300		115	80 - 120	P
Copper	ug/L	1500	1750		117	80 - 120	P
Lead	ug/L	5000	5500		110	80 - 120	P
Nickel	ug/L	2500	2800		112	80 - 120	P
Selenium	ug/L	10000	11100		111	80 - 120	P
Silver	ug/L	380	424		112	80 - 120	P
Zinc	ug/L	1000	1170		117	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q2236
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q2236 **SAS No.:** Q2236

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168335BS Mercury	ug/L	4.0	3.61		90	80 - 120	CV