

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

Continuing Calib Date: 06/07/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 02:31 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1268-1	(1)	7.60	7.61	7.51	7.71	0.01
Aroclor-1268-2	(2)	7.67	7.67	7.57	7.77	0.00
Aroclor-1268-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1268-4	(4)	8.16	8.17	8.07	8.27	0.01
Aroclor-1268-5	(5)	8.46	8.47	8.37	8.57	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.72	8.73	8.63	8.83	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

Continuing Calib Date: 06/07/2025 Initial Calibration Date(s): 05/14/2025 05/15/2025

Continuing Calib Time: 02:31 Initial Calibration Time(s): 17:19 01:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1268-1	(1)	7.57	7.57	7.47	7.67	0.00
Aroclor-1268-2	(2)	7.63	7.64	7.54	7.74	0.01
Aroclor-1268-3	(3)	7.84	7.84	7.74	7.94	0.00
Aroclor-1268-4	(4)	8.12	8.13	8.03	8.23	0.01
Aroclor-1268-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.67	8.67	8.57	8.77	0.00