

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

Continuing Calib Date: 06/06/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 17:18 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1242-1       | (1) | 5.66       | 5.66      | 5.56      | 5.76  | 0.00       |
| Aroclor-1242-2       | (2) | 5.68       | 5.68      | 5.58      | 5.78  | 0.00       |
| Aroclor-1242-3       | (3) | 5.74       | 5.74      | 5.64      | 5.84  | 0.00       |
| Aroclor-1242-4       | (4) | 5.84       | 5.84      | 5.74      | 5.94  | 0.00       |
| Aroclor-1242-5       | (5) | 6.57       | 6.57      | 6.47      | 6.67  | 0.00       |
| Tetrachloro-m-xylene |     | 4.50       | 4.51      | 4.41      | 4.61  | 0.01       |
| Decachlorobiphenyl   |     | 10.21      | 10.22     | 10.12     | 10.32 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

Continuing Calib Date: 06/06/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 17:18 Initial Calibration Time(s): 09:42 18:25

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1242-1       | (1) | 4.88       | 4.88      | 4.78      | 4.98 | 0.00       |
| Aroclor-1242-2       | (2) | 4.90       | 4.90      | 4.80      | 5.00 | 0.01       |
| Aroclor-1242-3       | (3) | 5.07       | 5.08      | 4.98      | 5.18 | 0.01       |
| Aroclor-1242-4       | (4) | 5.16       | 5.16      | 5.06      | 5.26 | 0.00       |
| Aroclor-1242-5       | (5) | 5.68       | 5.69      | 5.59      | 5.79 | 0.01       |
| Tetrachloro-m-xylene |     | 3.80       | 3.80      | 3.70      | 3.90 | 0.00       |
| Decachlorobiphenyl   |     | 8.81       | 8.83      | 8.73      | 8.93 | 0.02       |