

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO051525

Review By	yogesh	Review On	5/14/2025 3:39:03 PM
Supervise By	mohammad	Supervise On	5/16/2025 1:55:32 AM
SubDirectory	PO051525	HP Acquire Method	HP Processing Method PO051525

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO111056.D	14 May 2025 16:43		YP/AJ	Ok
2	I.BLK	I.BLK	PO111057.D	14 May 2025 17:01		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO111058.D	14 May 2025 17:19		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO111059.D	14 May 2025 17:38		YP/AJ	Ok,M
5	AR1660ICC500	AR1660ICC500	PO111060.D	14 May 2025 17:56		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO111061.D	14 May 2025 18:15		YP/AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PO111062.D	14 May 2025 18:33		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO111063.D	14 May 2025 18:51		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO111064.D	14 May 2025 19:10		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO111065.D	14 May 2025 19:28		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO111066.D	14 May 2025 19:46		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO111067.D	14 May 2025 20:05		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO111068.D	14 May 2025 20:23		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO111069.D	14 May 2025 20:41		YP/AJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PO111070.D	14 May 2025 21:00		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO111071.D	14 May 2025 21:18		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO111072.D	14 May 2025 21:37		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO111073.D	14 May 2025 21:54		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO051525

Review By	yogesh	Review On	5/14/2025 3:39:03 PM
Supervise By	mohammad	Supervise On	5/16/2025 1:55:32 AM
SubDirectory	PO051525	HP Acquire Method	HP Processing Method PO051525
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC050	AR1248ICC050	PO111074.D	14 May 2025 22:12		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO111075.D	14 May 2025 22:31		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO111076.D	14 May 2025 22:49		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO111077.D	14 May 2025 23:06		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO111078.D	14 May 2025 23:25		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO111079.D	14 May 2025 23:43		YP/AJ	Ok
25	AR1262ICC500	AR1262ICC500	PO111080.D	15 May 2025 00:01		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO111081.D	15 May 2025 00:19		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO111082.D	15 May 2025 00:37		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO111083.D	15 May 2025 00:56		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO111084.D	15 May 2025 01:14		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO111085.D	15 May 2025 01:31		YP/AJ	Ok
31	PO051525ICV500	ICVPO051525	PO111086.D	15 May 2025 01:50		YP/AJ	Ok
32	AR1242ICV500	ICVPO051525AR1242	PO111087.D	15 May 2025 02:26		YP/AJ	Ok
33	AR1248ICV500	ICVPO051525AR1248	PO111088.D	15 May 2025 03:03		YP/AJ	Ok
34	AR1254ICV500	ICVPO051525AR1254	PO111089.D	15 May 2025 03:40		YP/AJ	Ok
35	AR1268ICV500	ICVPO051525AR1268	PO111090.D	15 May 2025 04:17		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO060625

Review By	yogesh	Review On	6/6/2025 11:18:47 AM
Supervise By		Supervise On	
SubDirectory	PO060625	HP Acquire Method	HP Processing Method PO051525

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO111511.D	06 Jun 2025 08:59		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO111512.D	06 Jun 2025 09:17	2nd peak high in 1st column for AR1260	YP/AJ	Ok,NS
3	AR1242CCC500	AR1242CCC500	PO111513.D	06 Jun 2025 09:36		YP/AJ	Ok,NS
4	AR1248CCC500	AR1248CCC500	PO111514.D	06 Jun 2025 09:54		YP/AJ	Ok,NS
5	AR1254CCC500	AR1254CCC500	PO111515.D	06 Jun 2025 10:12		YP/AJ	Ok,NS
6	I.BLK	I.BLK	PO111516.D	06 Jun 2025 10:31		YP/AJ	Ok,NS
7	Q2228-01	TP08-MHI-WC	PO111517.D	06 Jun 2025 12:39		YP/AJ	Ok,NS
8	Q2235-02	WC-A2-08-C	PO111518.D	06 Jun 2025 12:58		YP/AJ	Ok,NS
9	Q2236-02	WC-A4-05A-C	PO111519.D	06 Jun 2025 13:16		YP/AJ	Ok,NS
10	Q2236-06	WC-A2-04-C	PO111520.D	06 Jun 2025 13:35		YP/AJ	Ok,NS
11	Q2236-10	WC-A2-05-C	PO111521.D	06 Jun 2025 13:53		YP/AJ	Ok,NS
12	Q2236-14	WC-A2-06-C	PO111522.D	06 Jun 2025 14:12		YP/AJ	Ok,NS
13	Q2236-18	WC-A2-07-C	PO111523.D	06 Jun 2025 14:30		YP/AJ	Ok,NS
14	Q2238-01	0528	PO111524.D	06 Jun 2025 14:49	AR1242 Hit,Need Dilution; Need 5X	YP/AJ	Dilution
15	Q2240-01	TP-3	PO111525.D	06 Jun 2025 15:07	AR1268 Hit,AR1268CCC missing	YP/AJ	Not Ok
16	Q2240-05	TP-2	PO111526.D	06 Jun 2025 15:25	AR1254+ AR1268 Hit, AR1268CCC missing	YP/AJ	Not Ok
17	AR1660CCC500	AR1660CCC500	PO111527.D	06 Jun 2025 16:39		YP/AJ	Ok,NS

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO060625

Review By	yogesh	Review On	6/6/2025 11:18:47 AM
Supervise By		Supervise On	
SubDirectory	PO060625	HP Acquire Method	HP Processing Method PO051525
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PO111528.D	06 Jun 2025 16:56		YP/AJ	Ok,NS
19	AR1248CCC500	AR1248CCC500	PO111529.D	06 Jun 2025 17:14		YP/AJ	Ok,NS
20	AR1254CCC500	AR1254CCC500	PO111530.D	06 Jun 2025 17:31		YP/AJ	Ok,NS
21	AR1268CCC500	AR1268CCC500	PO111531.D	06 Jun 2025 17:48	DCB high in both column	YP/AJ	Ok,NS
22	I.BLK	I.BLK	PO111532.D	06 Jun 2025 18:06		YP/AJ	Ok,NS
23	Q2240-05	TP-2	PO111533.D	06 Jun 2025 18:23	Surrogates low in all column , AR1260+ 1254 + 1268 Hit	YP/AJ	Not Ok
24	AR1660CCC500	AR1660CCC500	PO111534.D	07 Jun 2025 01:19		YP/AJ	Ok,NS
25	AR1242CCC500	AR1242CCC500	PO111535.D	07 Jun 2025 01:37		YP/AJ	Ok,NS
26	AR1248CCC500	AR1248CCC500	PO111536.D	07 Jun 2025 01:56		YP/AJ	Ok,NS
27	AR1254CCC500	AR1254CCC500	PO111537.D	07 Jun 2025 02:13		YP/AJ	Ok,NS
28	AR1268CCC500	AR1268CCC500	PO111538.D	07 Jun 2025 02:31	DCB high in both column	YP/AJ	Ok,NS
29	I.BLK	I.BLK	PO111539.D	07 Jun 2025 02:49		YP/AJ	Ok,NS

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP072162.D	19 May 2025 09:09		YP\AJ	Ok
2	I.BLK	I.BLK	PP072163.D	19 May 2025 09:26		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP072164.D	19 May 2025 09:42		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PP072165.D	19 May 2025 09:59		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP072166.D	19 May 2025 10:15		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP072167.D	19 May 2025 10:32		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PP072168.D	19 May 2025 10:48		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP072169.D	19 May 2025 11:05		YP\AJ	Ok
9	AR1232ICC1000	AR1232ICC1000	PP072170.D	19 May 2025 11:21		YP\AJ	Ok
10	AR1232ICC750	AR1232ICC750	PP072171.D	19 May 2025 11:38		YP\AJ	Ok
11	AR1232ICC500	AR1232ICC500	PP072172.D	19 May 2025 11:54		YP\AJ	Ok
12	AR1232ICC250	AR1232ICC250	PP072173.D	19 May 2025 12:10		YP\AJ	Ok
13	AR1232ICC050	AR1232ICC050	PP072174.D	19 May 2025 12:27		YP\AJ	Ok,M
14	AR1242ICC1000	AR1242ICC1000	PP072175.D	19 May 2025 12:43		YP\AJ	Ok
15	AR1242ICC750	AR1242ICC750	PP072176.D	19 May 2025 12:59		YP\AJ	Ok
16	AR1242ICC500	AR1242ICC500	PP072177.D	19 May 2025 13:15		YP\AJ	Ok
17	AR1242ICC250	AR1242ICC250	PP072178.D	19 May 2025 13:32		YP\AJ	Ok
18	AR1242ICC050	AR1242ICC050	PP072179.D	19 May 2025 13:48		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC1000	AR1248ICC1000	PP072180.D	19 May 2025 14:04		YPIAJ	Ok
20	AR1248ICC750	AR1248ICC750	PP072181.D	19 May 2025 14:20		YPIAJ	Ok
21	AR1248ICC500	AR1248ICC500	PP072182.D	19 May 2025 14:37		YPIAJ	Ok
22	AR1248ICC250	AR1248ICC250	PP072183.D	19 May 2025 14:53		YPIAJ	Ok
23	AR1248ICC050	AR1248ICC050	PP072184.D	19 May 2025 15:25		YPIAJ	Ok,M
24	AR1254ICC1000	AR1254ICC1000	PP072185.D	19 May 2025 15:42		YPIAJ	Ok
25	AR1254ICC750	AR1254ICC750	PP072186.D	19 May 2025 15:58		YPIAJ	Ok
26	AR1254ICC500	AR1254ICC500	PP072187.D	19 May 2025 16:14		YPIAJ	Ok
27	AR1254ICC250	AR1254ICC250	PP072188.D	19 May 2025 16:31		YPIAJ	Ok
28	AR1254ICC050	AR1254ICC050	PP072189.D	19 May 2025 16:47		YPIAJ	Ok,M
29	AR1262ICC500	AR1262ICC500	PP072190.D	19 May 2025 17:03		YPIAJ	Ok
30	AR1268ICC1000	AR1268ICC1000	PP072191.D	19 May 2025 17:20		YPIAJ	Ok
31	AR1268ICC750	AR1268ICC750	PP072192.D	19 May 2025 17:36		YPIAJ	Ok
32	AR1268ICC500	AR1268ICC500	PP072193.D	19 May 2025 17:53		YPIAJ	Ok
33	AR1268ICC250	AR1268ICC250	PP072194.D	19 May 2025 18:09		YPIAJ	Ok
34	AR1268ICC050	AR1268ICC050	PP072195.D	19 May 2025 18:25		YPIAJ	Ok,M
35	PP051925ICV500	ICVPP051925	PP072196.D	19 May 2025 18:42		YPIAJ	Ok
36	AR1232ICV500	ICVPP051925AR1232	PP072197.D	19 May 2025 19:14		YPIAJ	Ok
37	AR1242ICV500	ICVPP051925AR1242	PP072198.D	19 May 2025 19:47		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP051925

Review By	yogesh	Review On	5/19/2025 12:22:08 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:37:35 AM
SubDirectory	PP051925	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369 PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

38	AR1248ICV500	ICVPP051925AR1248	PP072199.D	19 May 2025 20:19		YPIAJ	Ok
39	AR1254ICV500	ICVPP051925AR1254	PP072200.D	19 May 2025 20:52		YPIAJ	Ok
40	AR1268ICV500	ICVPP051925AR1268	PP072201.D	19 May 2025 21:25		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP060625

Review By	yogesh	Review On	6/6/2025 11:16:52 AM
Supervise By		Supervise On	
SubDirectory	PP060625	HP Acquire Method	HP Processing Method PP051925

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP072682.D	06 Jun 2025 08:57		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP072683.D	06 Jun 2025 09:13		YPIAJ	Ok,NS
3	AR1242CCC500	AR1242CCC500	PP072684.D	06 Jun 2025 09:30		YPIAJ	Ok,NS
4	AR1248CCC500	AR1248CCC500	PP072685.D	06 Jun 2025 09:47		YPIAJ	Ok,NS
5	AR1254CCC500	AR1254CCC500	PP072686.D	06 Jun 2025 10:03		YPIAJ	Ok,NS
6	I.BLK	I.BLK	PP072687.D	06 Jun 2025 10:19		YPIAJ	Ok,NS
7	Q2214-01	RBR251425	PP072688.D	06 Jun 2025 10:37		YPIAJ	Ok,NS
8	PB168319BL	PB168319BL	PP072689.D	06 Jun 2025 12:37		YPIAJ	Ok,NS
9	PB168319BS	PB168319BS	PP072690.D	06 Jun 2025 12:54		YPIAJ	Ok,NS
10	Q2223-01	HR-01-06042025	PP072691.D	06 Jun 2025 13:10		YPIAJ	Ok,NS
11	Q2223-01MS	HR-01-06042025MS	PP072692.D	06 Jun 2025 13:27		YPIAJ	Ok,NS
12	Q2223-01MSD	HR-01-06042025MSD	PP072693.D	06 Jun 2025 13:43	AR1016 recovery fail	YPIAJ	Ok,NS
13	Q2223-03	HR-04-06042025	PP072694.D	06 Jun 2025 13:59		YPIAJ	Ok,NS
14	Q2224-01	EO-01-06042025	PP072695.D	06 Jun 2025 14:16		YPIAJ	Ok,NS
15	Q2225-01	SU-03-06042025	PP072696.D	06 Jun 2025 14:32		YPIAJ	Ok,NS
16	Q2227-01	TP07-MHH-WC	PP072697.D	06 Jun 2025 14:49		YPIAJ	Ok
17	AR1660CCC500	AR1660CCC500	PP072698.D	06 Jun 2025 17:01		YPIAJ	Ok,NS
18	AR1242CCC500	AR1242CCC500	PP072699.D	06 Jun 2025 17:18		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP060625

Review By	yogesh	Review On	6/6/2025 11:16:52 AM
Supervise By		Supervise On	
SubDirectory	PP060625	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP072700.D	06 Jun 2025 17:34		YPIAJ	Ok,NS
20	AR1254CCC500	AR1254CCC500	PP072701.D	06 Jun 2025 17:51		YPIAJ	Ok,NS
21	I.BLK	I.BLK	PP072702.D	06 Jun 2025 18:07		YPIAJ	Ok,NS
22	Q2240-09	TP-1	PP072703.D	06 Jun 2025 18:24	AR1254 Hit	YPIAJ	Ok,NS
23	Q2241-01	TP-N	PP072704.D	06 Jun 2025 18:40		YPIAJ	Ok,NS
24	Q2241-05	TP-S	PP072705.D	06 Jun 2025 18:56		YPIAJ	Ok
25	Q2242-01	TP09-MHJ	PP072706.D	06 Jun 2025 19:13		YPIAJ	Ok
26	Q2244-01	TP03-MHC	PP072707.D	06 Jun 2025 19:29		YPIAJ	Ok
27	PB168325BL	PB168325BL	PP072708.D	06 Jun 2025 19:46		YPIAJ	Ok
28	PB168325BS	PB168325BS	PP072709.D	06 Jun 2025 20:02		YPIAJ	Ok,NS
29	PB168325BSD	PB168325BSD	PP072710.D	06 Jun 2025 20:18	AR1016 recovery fail	YPIAJ	Not Ok
30	Q2198-05	B-202-GW01	PP072711.D	06 Jun 2025 20:35		YPIAJ	Ok
31	Q2216-02	3887	PP072712.D	06 Jun 2025 20:51		YPIAJ	Ok
32	AR1660CCC500	AR1660CCC500	PP072713.D	06 Jun 2025 21:40		YPIAJ	Ok,NS
33	AR1242CCC500	AR1242CCC500	PP072714.D	06 Jun 2025 22:30		YPIAJ	Ok,NS
34	AR1248CCC500	AR1248CCC500	PP072715.D	06 Jun 2025 22:46		YPIAJ	Ok,NS
35	AR1254CCC500	AR1254CCC500	PP072716.D	06 Jun 2025 23:03		YPIAJ	Ok,NS
36	I.BLK	I.BLK	PP072717.D	06 Jun 2025 23:19		YPIAJ	Ok,NS
37	Q2216-03	3888	PP072718.D	06 Jun 2025 23:35	TCMX high in 1st column	YPIAJ	Ok,NS

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP060625

Review By	yogesh	Review On	6/6/2025 11:16:52 AM
Supervise By		Supervise On	
SubDirectory	PP060625	HP Acquire Method	HP Processing Method PP051925
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q2216-04	3864	PP072719.D	06 Jun 2025 23:52	F Flag in TCMX in 2nd column	YPIAJ	Not Ok
39	Q2216-05	3865	PP072720.D	07 Jun 2025 00:08		YPIAJ	Ok,NS
40	Q2216-06	3851	PP072721.D	07 Jun 2025 00:25		YPIAJ	Ok,NS
41	Q2237-02	TW-WTS-10	PP072722.D	07 Jun 2025 00:41		YPIAJ	Ok,NS
42	Q2221-02	34900	PP072723.D	07 Jun 2025 00:58	AR1242 Hit	YPIAJ	Ok,NS
43	Q2221-03	34867	PP072724.D	07 Jun 2025 01:14		YPIAJ	Ok
44	Q2218-03	3309	PP072725.D	07 Jun 2025 01:31		YPIAJ	Ok,NS
45	AR1660CCC500	AR1660CCC500	PP072726.D	07 Jun 2025 02:36		YPIAJ	Ok,NS
46	AR1242CCC500	AR1242CCC500	PP072727.D	07 Jun 2025 03:25		YPIAJ	Ok,NS
47	AR1248CCC500	AR1248CCC500	PP072728.D	07 Jun 2025 03:42		YPIAJ	Ok,NS
48	AR1254CCC500	AR1254CCC500	PP072729.D	07 Jun 2025 03:58		YPIAJ	Ok,NS
49	I.BLK	I.BLK	PP072730.D	07 Jun 2025 04:15		YPIAJ	Ok,NS

M : Manual Integration