

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL02 Date Analyzed: 06/06/2025

Lab Sample No.: HSTDCCC750 Data File : PS030547.D Time Analyzed: 23:12

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.394	9.304	9.504	737.410	712.500	3.5
2,4-D	8.503	8.412	8.612	703.020	705.000	-0.3
2,4-DCAA	7.362	7.269	7.469	735.100	750.000	-2.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL02 Date Analyzed: 06/06/2025

Lab Sample No.: HSTDCCC750 Data File : PS030547.D Time Analyzed: 23:12

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.937	9.842	10.042	729.380	712.500	2.4
2,4-D	9.030	8.934	9.134	716.460	705.000	1.6
2,4-DCAA	7.771	7.674	7.874	730.020	750.000	-2.7