

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL03 Date Analyzed: 06/07/2025

Lab Sample No.: HSTDCCC750 Data File : PS030560.D Time Analyzed: 04:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.393	9.304	9.504	746.690	712.500	4.8
2,4-D	8.502	8.412	8.612	721.770	705.000	2.4
2,4-DCAA	7.362	7.269	7.469	759.650	750.000	1.3

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL03 Date Analyzed: 06/07/2025

Lab Sample No.: HSTDCCC750 Data File : PS030560.D Time Analyzed: 04:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.938	9.842	10.042	736.020	712.500	3.3
2,4-D	9.030	8.934	9.134	754.140	705.000	7.0
2,4-DCAA	7.771	7.674	7.874	744.090	750.000	-0.8