

## Report of Analysis

Client:	ENTACT		Date Collected:	06/04/25	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	06/04/25	
Client Sample ID:	PIBLK-PS030475.D		SDG No.:	Q2236	
Lab Sample ID:	I.BLK-PS030475.D		Matrix:	TCLP	
Analytical Method:	8151A		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Herbicide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS030475.D	1		06/04/25	PS060425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
94-75-7	2,4-D	0.00092	U	0.00092	0.0020	mg/L
93-72-1	2,4,5-TP (Silvex)	0.00078	U	0.00078	0.0020	mg/L
<b>SURROGATES</b>						
19719-28-9	2,4-DCAA	445		70 (61) - 130 (136)	89%	SPK: 500

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit