

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

Continuing Calib Date: 06/10/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 14:49 Initial Calibration Time(s): 11:31 12:25

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.07	9.08	8.98	9.18	0.01
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endrin	6.57	6.58	6.48	6.68	0.01
Methoxychlor	7.49	7.50	7.40	7.60	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

Continuing Calib Date: 06/10/2025 Initial Calibration Date(s): 05/19/2025 05/19/2025

Continuing Calib Time: 14:49 Initial Calibration Time(s): 11:31 12:25

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	8.07	8.08	7.98	8.18	0.01
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.08	4.09	3.99	4.19	0.01
Heptachlor epoxide	4.87	4.88	4.78	4.98	0.01
Endrin	5.79	5.79	5.69	5.89	0.00
Methoxychlor	6.76	6.76	6.66	6.86	0.00