

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/09/2025

Lab Sample No.: PSTDCCC050 Data File : PD088822.D Time Analyzed: 10:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.073	8.977	9.177	52.790	50.000	5.6
Endrin	6.575	6.477	6.677	54.680	50.000	9.4
gamma-BHC (Lindane)	4.330	4.232	4.432	54.180	50.000	8.4
Heptachlor	4.929	4.832	5.032	55.030	50.000	10.1
Heptachlor epoxide	5.691	5.593	5.793	53.970	50.000	7.9
Methoxychlor	7.494	7.395	7.595	54.010	50.000	8.0
Tetrachloro-m-xylene	3.549	3.451	3.651	57.870	50.000	15.7

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/09/2025

Lab Sample No.: PSTDCCC050 Data File : PD088822.D Time Analyzed: 10:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.072	7.977	8.177	48.910	50.000	-2.2
Endrin	5.789	5.692	5.892	46.190	50.000	-7.6
gamma-BHC (Lindane)	3.728	3.631	3.831	47.180	50.000	-5.6
Heptachlor	4.082	3.985	4.185	46.540	50.000	-6.9
Heptachlor epoxide	4.872	4.775	4.975	46.860	50.000	-6.3
Methoxychlor	6.754	6.658	6.858	45.820	50.000	-8.4
Tetrachloro-m-xylene	2.879	2.782	2.982	47.210	50.000	-5.6