

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q2236 SAS No.: Q2236 SDG NO.: Q2236

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No. (PEM): PEM - PD088584.D Date Analyzed: 05/19/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:04

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.076	8.980	9.180	21.510	20.000	7.6
Tetrachloro-m-xylene	3.550	3.500	3.600	19.960	20.000	-0.2
alpha-BHC	3.999	3.950	4.050	9.120	10.000	-8.8
beta-BHC	4.516	4.470	4.570	10.260	10.000	2.6
gamma-BHC (Lindane)	4.331	4.280	4.380	9.460	10.000	-5.4
Endrin	6.576	6.510	6.650	50.860	50.000	1.7
4,4'-DDT	7.023	6.950	7.090	101.210	100.000	1.2
Methoxychlor	7.495	7.420	7.570	234.710	250.000	-6.1

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No. (PEM): PEM - PD088584.D Date Analyzed: 05/19/2025

Lab Sample No.(PEM): PEM Time Analyzed: 11:04

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	8.076	7.980	8.180	20.960	20.000	4.8
Tetrachloro-m-xylene	2.881	2.830	2.930	20.340	20.000	1.7
alpha-BHC	3.394	3.340	3.440	10.570	10.000	5.7
beta-BHC	4.027	3.980	4.080	10.990	10.000	9.9
gamma-BHC (Lindane)	3.731	3.680	3.780	10.670	10.000	6.7
Endrin	5.792	5.720	5.860	48.470	50.000	-3.1
4,4'-DDT	6.186	6.120	6.260	95.100	100.000	-4.9
Methoxychlor	6.757	6.690	6.830	195.630	250.000	-21.7